



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/24/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:55
In Date: 10/23/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/24/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133070

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4482-01	MYE4D0	1	1.16	8.58	9.74	9.45	96.6	
P4482-02	MYE4D1	2	1.15	8.66	9.81	9.7	98.7	
P4482-03	MYE4D2	3	1.17	8.57	9.74	9.42	96.3	
P4482-04	MYE4D3	4	1.16	8.60	9.76	9.68	99.1	
P4482-05	MYE4D4	5	1.17	8.36	9.53	9.15	95.5	
P4482-06	MYE4D5	6	1.17	8.71	9.88	9.75	98.5	
P4482-07	MYE4D6	7	1.17	8.71	9.88	9.47	95.3	
P4482-08	MYE4D7	8	1.16	8.37	9.53	9.3	97.3	
P4482-09	MYE4D8	9	1.16	8.42	9.58	8.82	91.0	
P4482-10	MYE4D9	10	1.15	8.67	9.82	9.04	91.0	
P4482-11	MYE4E0	11	1.14	8.55	9.69	9.36	96.1	
P4482-12	MYE4E1	12	1.14	8.58	9.72	9.44	96.7	
P4482-13	MYE4E2	13	1.17	8.75	9.92	9.3	92.9	
P4482-14	MYE4E3	14	1.18	8.35	9.53	9.35	97.8	
P4482-15	MYE4E3D	15	1.18	8.35	9.53	9.35	97.8	
P4482-16	MYE4E3S	16	1.18	8.35	9.53	9.35	97.8	
P4482-17	MYE4E4	17	1.16	8.75	9.91	9.69	97.5	
P4482-18	MYE4E5	18	1.16	8.63	9.79	9.44	95.9	
P4482-19	MYE4E6	19	1.15	8.59	9.74	9.17	93.4	
P4482-20	MYE4E7	20	1.16	8.54	9.7	9.22	94.4	
P4482-21	MYE4E8	21	1.15	8.48	9.63	9.08	93.5	
P4482-22	MYE4E9	22	1.14	8.58	9.72	8.83	89.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$