


SFAM01.1

INSTRUMENT ID: CV2



A= 0.0000e+000

B= 2.4467e-004 *slope*

C= 4.4897e-003 *4.4897e-003*

Rho= 0.9999630

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.024	0.024	78	0.000	78					
0.05	0.050					442	0				
0.20	0.200	0.211	0.011	844	0.0 %	844					
2.50	2.500	2.471	-0.029	10081	0.0 %	10081					
5.00	5.000	4.952	-0.048	20223	0.0 %	20223					
7.50	7.500	7.547	0.047	30829	0.0 %	30829					
10.0	10.000	9.995	-0.005	40832	0.0 %	40832					

0/00

LB133074

INMSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	78	-		0 SFAM01.1	PPB	10/23/2024 14:52	Std
	0.2 S01	844	-		0.2 SFAM01.1	PPB	10/23/2024 14:55	Std
	2.5 S02	10081	-		2.5 SFAM01.1	PPB	10/23/2024 14:57	Std
	5 S03	20223	-		5 SFAM01.1	PPB	10/23/2024 14:59	Std
	7.5 S04	30829	-		7.5 SFAM01.1	PPB	10/23/2024 15:02	Std
	10 S05	40832	-		10 SFAM01.1	PPB	10/23/2024 15:09	Std
ICV049	ICV049	14895	3.6488 -		SFAM01.1	PPB	10/23/2024 15:12	SMPL
ICB049	ICB049	18	0.0089 -		SFAM01.1	PPB	10/23/2024 15:17	SMPL
CCV060	CCV060	20438	5.005 -		SFAM01.1	PPB	10/23/2024 15:19	SMPL
CCB060	CCB060	34	0.0128 -		SFAM01.1	PPB	10/23/2024 15:24	SMPL
PB164357BL	PBS357	47	0.016 -		SFAM01.1	PPB	10/23/2024 15:26	SMPL
P3391-03	MDL-SOIL	234	0.0617 -		SFAM01.1	PPB	10/23/2024 15:29	SMPL
CCV061	CCV061	20533	5.0282 -		SFAM01.1	PPB	10/23/2024 15:31	SMPL
CCB061	CCB061	32	0.0123 -		SFAM01.1	PPB	10/23/2024 15:36	SMPL