



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/24/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:50
In Date: 10/23/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:17
Out Date: 10/24/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133076

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4499-01	MBHCJ7	1	1.16	8.48	9.64	8.32	84.4	
P4499-02	MBHCJ8	2	1.15	8.40	9.55	8.27	84.8	
P4499-03	MBHCJ9	3	1.15	8.60	9.75	7.95	79.1	
P4499-04	MBHCK0	4	1.16	8.51	9.67	7.99	80.3	
P4499-05	MBHCK1	5	1.12	8.76	9.88	8.09	79.6	
P4499-06	MBHCK2	6	1.12	8.77	9.89	7.94	77.8	
P4499-07	MBHCK8	7	1.12	8.77	9.89	9.32	93.5	
P4499-08	MBHCK9	8	1.15	8.39	9.54	8.85	91.8	
P4499-09	MBHCL0	9	1.17	8.53	9.7	9.04	92.3	
P4499-10	MBHCL1	10	1.15	8.84	9.99	9.25	91.6	
P4499-11	MBHCL2	11	1.15	8.42	9.57	8.59	88.4	
P4499-12	MBHCY0	12	1.18	8.46	9.64	7.27	72.0	
P4499-13	MBHCL8	13	1.14	8.66	9.8	8.36	83.4	
P4499-14	MBHCL9	14	1.18	8.48	9.66	8.38	84.9	
P4499-15	MBHCM0	15	1.19	8.66	9.85	7.89	77.4	
P4499-16	MBHCM1	16	1.19	8.50	9.69	8.76	89.1	
P4499-17	MBHCM2	17	1.15	8.73	9.88	8.14	80.1	
P4499-18	MBHCM8	18	1.15	8.83	9.98	8.39	82.0	
P4499-19	MBHCM9	19	1.16	8.50	9.66	8.85	90.5	
P4499-20	MBHCN0	20	1.14	8.85	9.99	8.72	85.6	
P4499-21	MBHCN0D	21	1.14	8.85	9.99	8.72	85.6	
P4499-22	MBHCN0S	22	1.14	8.85	9.99	8.72	85.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$