

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133083

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	10/23/2024	10:40
2	P4488-01		1	11	100.00	>212.0	>212.0	10/23/2024	11:10
3	P4488-01DUP		1	11	100.00	>212.0	>212.0	10/23/2024	11:40
4	P4488-02		1	4	7.00	44.6	44.3	10/23/2024	12:10
5	P4488-09		1	6	64.00	147.2	146.9	10/23/2024	12:40
6	P4488-10		1	6	67.00	152.6	152.3	10/23/2024	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133683

WorkList Name : FP-10-22

WorkList ID : 184672

Department : Wet-Chemistry

Date : 10-22-2024 16:36:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4488-01	OIL-1	Solid	Flash Point	Cool 4 deg C	PSEG03	K61	10/22/2024	1010B
P4488-02	1017-B	Solid	Flash Point	Cool 4 deg C	PSEG03	K61	10/22/2024	1010B
P4488-09	HCC-1	Solid	Flash Point	Cool 4 deg C	PSEG03	K61	10/22/2024	1010B
P4488-10	HCC-2	Solid	Flash Point	Cool 4 deg C	PSEG03	K61	10/22/2024	1010B

Date/Time 10/23/2024 10:30
 Raw Sample Received by: RM
 Raw Sample Relinquished by: JPL Cell 1

Date/Time 10/23/2024 13:50
 Raw Sample Received by: JPL Cell 1
 Raw Sample Relinquished by: RPL Cell 1