

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133084

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	10/23/2024	08:20
2	P4453-01		1	8	66.00	150.8	150.5	10/23/2024	08:50
3	P4485-03		1	11	100.00	>212.0	>212.0	10/23/2024	09:20
4	P4485-03DUP		1	11	100.00	>212.0	>212.0	10/23/2024	09:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16133084

WorkList Name : fp-1021

WorkList ID : 184619

Department : Wet-Chemistry

Date : 10-21-2024 08:19:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4453-01	34838	Water	Flash Point	Cool 4 deg C	PSEG03	K51	10/18/2024	1010B
P4485-03	D20241001-02-03	Water	Flash Point	Cool 4 deg C	PSEG03	K62	10/22/2024	1010B

Date/Time 10/23/2024 8:10
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: JH CWC

Date/Time 10/23/2024 10:25
Raw Sample Received by: JH CWC
Raw Sample Relinquished by: RM CWC