

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133090

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/24/2024	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	10/24/2024	08:11
3	CAL3	1	Water	NA	NA	20.2	10.02	10/24/2024	08:12
4	ICV	1	Water	NA	NA	20.2	7.00	10/24/2024	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	10/24/2024	08:22
6	P4511-01	1	Water	NA	NA	23.2	7.82	10/24/2024	09:20
7	P4511-01DUP	1	Water	NA	NA	23.3	7.83	10/24/2024	09:21
8	P4512-01	1	Water	NA	NA	22.2	7.73	10/24/2024	09:22
9	P4512-02	1	Water	NA	NA	23.4	7.53	10/24/2024	09:25
10	P4513-04	1	Water	NA	NA	23.4	8.52	10/24/2024	09:30
11	CCV2	1	Water	NA	NA	20.2	12.02	10/24/2024	09:32

WORKLIST(Hardcopy Internal Chain)

133090

WorkList Name : ph p4509

WorkList ID : 184716

Department : Wet-Chemistry

Date : 10-24-2024 07:50:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4511-01	266	Water	pH	Cool 4 deg C	PSEG03	K61	10/23/2024	9040C
P4512-01	3140	Water	pH	Cool 4 deg C	PSEG03	K61	10/23/2024	9040C
P4512-02	3149	Water	pH	Cool 4 deg C	PSEG03	K61	10/23/2024	9040C
P4513-04	D3682	Water	pH	Cool 4 deg C	PSEG03	K61	10/23/2024	9040C

Date/Time 10/24/24 08:50

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 10/24/24 10:15

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]