


LB133099

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

43125



Linear

A= 0.0000e+000

B= 2.2969e-004 slope

C= 2.6915e-003 y-intercept

Rho= 0.9996610

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std ID
0.00	0.000	0.004	0.004	7	0.000	7					0.00
0.05	0.050										0.05
0.20	0.200	0.203	0.003	872	0.0 %	872					0.20
2.50	2.500	2.509	0.009	10914	0.0 %	10914					2.50
5.00	5.000	4.889	-0.111	21276	0.0 %	21276					5.00
7.50	7.500	7.686	0.186	33451	0.0 %	33451					7.50
10.0	10.000	9.908	-0.092	43125	0.0 %	43125					10.0

Std ID
0.00
0.05
0.20
2.50
5.00
7.50
10.0

LB133099

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	7	-		0 SFAM01.1	PPB	10/24/2024 11:57	Std
	0.2 S01	872	-		0.2 SFAM01.1	PPB	10/24/2024 12:00	Std
	2.5 S02	10914	-		2.5 SFAM01.1	PPB	10/24/2024 12:02	Std
	5 S03	21276	-		5 SFAM01.1	PPB	10/24/2024 12:04	Std
	7.5 S04	33451	-		7.5 SFAM01.1	PPB	10/24/2024 12:07	Std
	10 S05	43125	-		10 SFAM01.1	PPB	10/24/2024 12:12	Std
ICV051	ICV051	16815	3.8649 -		SFAM01.1	PPB	10/24/2024 12:15	SMPL
ICB051	ICB051	-72	-0.0138 -		SFAM01.1	PPB	10/24/2024 12:19	SMPL
CCV066	CCV066	21827	5.016 -		SFAM01.1	PPB	10/24/2024 12:22	SMPL
CCB066	CCB066	-53	-0.0095 -		SFAM01.1	PPB	10/24/2024 12:27	SMPL
PB164380BL	PBS380	-48	-0.0083 -		SFAM01.1	PPB	10/24/2024 12:29	SMPL
P3391-04	MDL-SOIL	191	0.0466 -		SFAM01.1	PPB	10/24/2024 12:31	SMPL
CCV067	CCV067	21635	4.9719 -		SFAM01.1	PPB	10/24/2024 12:33	SMPL
CCB067	CCB067	-45	-0.0076 -		SFAM01.1	PPB	10/24/2024 12:38	SMPL