



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/25/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 10/24/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:58
Out Date: 10/25/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133100

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4501-01	MBHCS8	1	1.14	8.40	9.54	8.64	89.3	
P4501-02	MBHCS9	2	1.18	8.62	9.8	8.94	90.0	
P4501-03	MBHCT0	3	1.13	8.73	9.86	8.99	90.0	
P4501-04	MBHCT1	4	1.19	8.58	9.77	8.63	86.7	
P4501-05	MBHCT2	5	1.19	8.52	9.71	8.3	83.5	
P4501-06	MBHCT3	6	1.15	8.82	9.97	8.35	81.6	
P4501-07	MBHCT4	7	1.15	8.82	9.97	8.31	81.2	
P4501-08	MBHCT5	8	1.2	8.72	9.92	8.37	82.2	
P4501-09	MBHCT6	9	1.15	8.40	9.55	7.84	79.6	
P4501-10	MBHCT7	10	1.15	8.79	9.94	8.22	80.4	
P4501-11	MBHCT8	11	1.16	8.70	9.86	8.64	86.0	
P4501-12	MBHCT9	12	1.19	8.43	9.62	8.04	81.3	
P4501-13	MBHCW0	13	1.14	8.64	9.78	8.17	81.4	
P4501-14	MBHCW0D	14	1.14	8.64	9.78	8.17	81.4	
P4501-15	MBHCW0S	15	1.14	8.64	9.78	8.17	81.4	
P4501-16	MBHCW1	16	1.16	8.64	9.8	8.05	79.7	
P4501-17	MBHCW2	17	1.1	8.77	9.87	7.99	78.6	
P4501-18	MBHCW3	18	1.18	8.47	9.65	7.65	76.4	
P4501-19	MBHCW4	19	1.12	8.65	9.77	7.65	75.5	
P4501-20	MBHCW5	20	1.16	8.62	9.78	7.93	78.5	
P4501-21	MBHCW6	21	1.13	8.60	9.73	7.45	73.5	
P4501-22	MBHCW7	22	1.19	8.50	9.69	8.06	80.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$