



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/25/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:50
In Date: 10/24/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/25/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133103

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4502-01	MBHCY1	1	1.15	8.49	9.64	8.03	81.0	
P4502-02	MBHCY2	2	1.15	8.79	9.94	7.47	71.9	
P4502-03	MBHCY3	3	1.15	8.82	9.97	8.71	85.7	
P4502-04	MBHCY4	4	1.14	8.46	9.6	5.92	56.5	
P4502-05	MBHCY8	5	1.15	8.35	9.5	7.75	79.0	
P4502-06	MBHCY9	6	1.19	8.53	9.72	8.13	81.4	
P4502-07	MBHCZ0	7	1.19	8.53	9.72	7.47	73.6	
P4502-08	MBHCZ1	8	1.14	8.76	9.9	8.36	82.4	
P4502-09	MBHCZ2	9	1.14	8.81	9.95	8.08	78.8	
P4502-10	MBHCZ3	10	1.16	8.75	9.91	8.72	86.4	
P4502-11	MBHCX4	11	1.18	8.54	9.72	8.04	80.3	
P4502-12	MBHCX7	12	1.19	8.41	9.6	8.14	82.6	
P4502-13	MBHCX8	13	1.15	8.40	9.55	7.79	79.0	
P4502-14	MBHCX9	14	1.15	8.38	9.53	6.68	66.0	
P4502-15	MBHDO0	15	1.19	8.42	9.61	7.99	80.8	
P4502-16	MBHDO1	16	1.19	8.50	9.69	8.09	81.2	
P4502-17	MBHDO2	17	1.17	8.40	9.57	7.74	78.2	
P4502-18	MBHDO3	18	1.18	8.52	9.7	7.75	77.1	
P4502-19	MBHDO4	19	1.17	8.80	9.97	8.15	79.3	
P4502-20	MBHZC9	20	1.16	8.82	9.98	8.45	82.7	
P4502-21	MBHZC9D	21	1.16	8.82	9.98	8.45	82.7	
P4502-22	MBHZC9S	22	1.16	8.82	9.98	8.45	82.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$