


LB133109

7470A

INSTRUMENT NO: CV12



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.013	0.013	645	0.000	645	0				1
0.2	0.200	0.221	0.021	1428	0.0 %	1428					11
2.5	2.500	2.442	-0.058	9793	0.0 %	9793					12
5.0	5.000	4.997	-0.003	19412	0.0 %	19412					0
7.5	7.500	7.551	0.051	29027	0.0 %	29027					0
10.0	10.000	9.977	-0.023	38163	0.0 %	38163					0

LB133109 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Oct 2024 13:45:52

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	645	-	0.0000	7470A	PPB	24 Oct 2024 14:33:59	S	Std
0.2 - 1	50.2	1428	-	0.2000	7470A	PPB	24 Oct 2024 14:36:16	S	Std
2.5 - 1	52.5	9793	-	2.5000	7470A	PPB	24 Oct 2024 14:38:33	S	Std
5.0 - 1	55	19412	-	5.0000	7470A	PPB	24 Oct 2024 14:40:49	S	Std
7.5 - 1	57.5	29027	-	7.5000	7470A	PPB	24 Oct 2024 14:45:59	S	Std
10.0 - 1	511	38163	-	10.0000	7470A	PPB	24 Oct 2024 14:48:14	S	Std
ICV52 - 1	ICV52	15597	3.9837	-	7470A	PPB	24 Oct 2024 15:02:37	U	SMPL
ICB52 - 1	ICB52	405	-0.0511	-	7470A	PPB	24 Oct 2024 15:04:51	U	SMPL
CCV68 - 1	CCV68	18983	4.8830	-	7470A	PPB	24 Oct 2024 15:07:08	U	SMPL
CCB68 - 1	CCB68	284	-0.0832	-	7470A	PPB	24 Oct 2024 15:09:23	U	SMPL
CRA - 1	CRA	1356	0.2015	-	7470A	PPB	24 Oct 2024 15:11:40	U	SMPL
HighStd - 1	HighStd	37442	9.7855	-	7470A	PPB	24 Oct 2024 15:13:56	U	SMPL
ChkStd - 1	ChkStd	25769	6.6853	-	7470A	PPB	24 Oct 2024 15:16:10	U	SMPL
PB164370BL - 1	PBW	211	-0.1026	-	7470A	PPB	24 Oct 2024 15:18:29	U	SMPL
PB164370BS - 1	LCSW	15884	4.0599	-	7470A	PPB	24 Oct 2024 15:20:47	U	SMPL
P4485-02 - 1	D20241001-01-04	370	-0.0604	-	7470A	PPB	24 Oct 2024 15:23:03	U	SMPL
P4485-02DUP - 1	D20241001-01-04DUP	562	-0.0094	-	7470A	PPB	24 Oct 2024 15:25:21	U	SMPL
P4485-02MS - 1	D20241001-01-04MS	15549	3.9710	-	7470A	PPB	24 Oct 2024 15:30:14	U	SMPL
P4485-02MSD - 1	D20241001-01-04MSD	15343	3.9163	-	7470A	PPB	24 Oct 2024 15:32:31	U	SMPL
P4487-04 - 1	BP-B5	401	-0.0521	-	7470A	PPB	24 Oct 2024 15:34:47	U	SMPL
CCV69 - 1	CCV69	19076	4.9077	-	7470A	PPB	24 Oct 2024 15:37:04	U	SMPL
CCB69 - 1	CCB69	366	-0.0614	-	7470A	PPB	24 Oct 2024 15:39:20	U	SMPL
P4487-08 - 1	BP-B27	454	-0.0380	-	7470A	PPB	24 Oct 2024 15:41:37	U	SMPL
P4488-09 - 1	HCC-1	1386	0.2095	-	7470A	PPB	24 Oct 2024 15:43:52	U	SMPL
P4488-10 - 1	HCC-2	845	0.0658	-	7470A	PPB	24 Oct 2024 15:46:08	U	SMPL
P4508-04 - 1	TP-3	-162	-0.2016	-	7470A	PPB	24 Oct 2024 15:48:24	U	SMPL
P4508-08 - 1	BP-F23	-86	-0.1815	-	7470A	PPB	24 Oct 2024 15:50:40	U	SMPL
P4508-12 - 1	BP-F22	-125	-0.1918	-	7470A	PPB	24 Oct 2024 15:52:56	U	SMPL
P4511-02 - 1	267	16	-0.1544	-	7470A	PPB	24 Oct 2024 15:55:12	U	SMPL
P4513-01 - 1	D3683	450	-0.0391	-	7470A	PPB	24 Oct 2024 15:57:29	U	SMPL
P4513-02 - 1	D3694	384	-0.0566	-	7470A	PPB	24 Oct 2024 15:59:45	U	SMPL
P4513-03 - 1	D3695	315	-0.0750	-	7470A	PPB	24 Oct 2024 16:02:01	U	SMPL
CCV70 - 1	CCV70	19781	5.0949	-	7470A	PPB	24 Oct 2024 16:25:49	U	SMPL
CCB70 - 1	CCB70	147	-0.1196	-	7470A	PPB	24 Oct 2024 16:28:06	U	SMPL
P4515-02 - 1	CHVB0783	766	0.0448	-	7470A	PPB	24 Oct 2024 16:30:25	U	SMPL
P4516-01 - 1	72-11986	27	-0.1515	-	7470A	PPB	24 Oct 2024 16:32:41	U	SMPL
PB164371BL - 1	PBW	412	-0.0492	-	7470A	PPB	24 Oct 2024 16:34:58	U	SMPL
PB164371BS - 1	LCSW	14531	3.7006	-	7470A	PPB	24 Oct 2024 16:39:48	U	SMPL
P4511-01 - 1	266	7	-0.1568	-	7470A	PPB	24 Oct 2024 16:42:04	U	SMPL
P4512-01 - 1	3140	610	0.0034	-	7470A	PPB	24 Oct 2024 16:44:22	U	SMPL
P4512-02 - 1	3149	703	0.0281	-	7470A	PPB	24 Oct 2024 16:46:38	U	SMPL
P4513-04 - 1	D3682	491	-0.0282	-	7470A	PPB	24 Oct 2024 16:48:55	U	SMPL
P4513-04DUP - 1	D3682DUP	507	-0.0240	-	7470A	PPB	24 Oct 2024 16:51:11	U	SMPL
P4513-04MS - 1	D3682MS	9157	2.2733	-	7470A	PPB	24 Oct 2024 16:53:27	U	SMPL
CCV71 - 1	CCV71	18422	4.7340	-	7470A	PPB	24 Oct 2024 16:55:42	U	SMPL
CCB71 - 1	CCB71	191	-0.1079	-	7470A	PPB	24 Oct 2024 16:58:00	U	SMPL
P4513-04MSD - 1	D3682MSD	8425	2.0789	-	7470A	PPB	24 Oct 2024 17:00:19	U	SMPL
PB164335TB - 1	PB164336TB	493	-0.0277	-	7470A	PPB	24 Oct 2024 17:05:15	U	SMPL
PB164336TB - 1	PB164336TB	484	-0.0301	-	7470A	PPB	24 Oct 2024 17:07:30	U	SMPL
P4485-02LX5 - 1		190	-0.1082	-	7470A	PPB	24 Oct 2024 17:09:45	U	SMPL
P4485-02A - 1		13978	3.5537	-	7470A	PPB	24 Oct 2024 17:12:01	U	SMPL
P4513-04LX5 - 1		44	-0.1469	-	7470A	PPB	24 Oct 2024 17:14:17	U	SMPL
P4513-04A - 1		17933	4.6041	-	7470A	PPB	24 Oct 2024 17:16:34	U	SMPL
CCV72 - 1	CCV72	19583	5.0424	-	7470A	PPB	24 Oct 2024 17:24:10	U	SMPL
CCB72 - 1	CCB72	192	-0.1076	-	7470A	PPB	24 Oct 2024 17:26:29	U	SMPL