

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB133111

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/24/2024	15:25
2	CAL2	1	Water	NA	NA	20.2	7.00	10/24/2024	15:26
3	CAL3	1	Water	NA	NA	20.3	10.02	10/24/2024	15:30
4	ICV	1	Water	NA	NA	20.3	7.00	10/24/2024	15:35
5	CCV1	1	Water	NA	NA	20.2	2.01	10/24/2024	15:37
6	P4545-02	1	Solid	20.02	20	21.4	7.62	10/24/2024	15:42
7	P4545-02DUP	1	Solid	20.03	20	21.5	7.63	10/24/2024	15:43
8	P4547-04	1	Solid	20.04	20	21.0	6.18	10/24/2024	15:47
9	P4547-08	1	Solid	20.02	20	21.4	8.13	10/24/2024	15:50
10	CCV2	1	Water	NA	NA	20.3	12.02	10/24/2024	16:00

WORKLIST(Hardcopy Internal Chain)

3133111

WorkList Name : corrosivity p4547 WorkList ID : 184745 Department : Wet-Chemistry Date : 10-24-2024 15:17:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4545-02	VNU-215	Solid	Corrosivity	Cool 4 deg C	PSEG03	K61	10/24/2024	9045D
P4547-04	BP-F-21	Solid	Corrosivity	Cool 4 deg C	PSEG03	K51	10/24/2024	9045D
P4547-08	BP-F-20	Solid	Corrosivity	Cool 4 deg C	PSEG03	K51	10/24/2024	9045D

Date/Time 10/24/24 15120
Raw Sample Received by: JH woc
Raw Sample Relinquished by: JT (S-M)

Date/Time 10/24/24 18:00
Raw Sample Received by: JT (S-M)
Raw Sample Relinquished by: JH woc