

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133113

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/24/2024	15:25
2	CAL2	1	Water	NA	NA	20.2	7.00	10/24/2024	15:26
3	CAL3	1	Water	NA	NA	20.3	10.02	10/24/2024	15:30
4	ICV	1	Water	NA	NA	20.3	7.00	10/24/2024	15:35
5	CCV1	1	Water	NA	NA	20.1	2.02	10/24/2024	16:10
6	P4546-01	1	Water	NA	NA	20.8	5.52	10/24/2024	16:15
7	P4546-01DUP	1	Water	NA	NA	20.9	5.53	10/24/2024	16:16
8	P4546-02	1	Water	NA	NA	21.1	5.19	10/24/2024	16:20
9	CCV2	1	Water	NA	NA	20.3	12.02	10/24/2024	16:22

WORKLIST(Hardcopy Internal Chain)

133113

WorkList Name : ph p4546

WorkList ID : 184744

Department : Wet-Chemistry

Date : 10-24-2024 15:16:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4546-01	34839	Water	pH	Cool 4 deg C	PSEG03	K61	10/24/2024	9040C
P4546-02	34840	Water	pH	Cool 4 deg C	PSEG03	K61	10/24/2024	9040C

Date/Time 10/24/24 15:20

Raw Sample Received by: JH Coors

Raw Sample Relinquished by: JT (SM)

Date/Time 10/24/24

Raw Sample Received by: JT (SM)

Raw Sample Relinquished by: JH Coors