

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133116

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	10/24/2024	15:45
2	P4513-01		1	4	6.00	42.8	42.5	10/24/2024	16:15
3	P4513-01DUP		1	4	6.00	42.8	42.5	10/24/2024	16:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp s-10-24

WorkList ID : 184728

Department : Wet-Chemistry

Date : 10-24-2024 09:02:03

66133116

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4513-01	D3683	Solid	Flash Point	Cool 4 deg C	PSEG03	K61	10/23/2024	1010B

Date/Time 10/24/2024 15:35
Raw Sample Received by: RY CWS
Raw Sample Relinquished by: JH WDC1

Date/Time 10/24/2024 17:20
Raw Sample Received by: JH WDC1
Raw Sample Relinquished by: RY CWS