

Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB133121

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP110410
Turbidity Calibration std, 20NTU	WP110405
Turbidity Calibration std, 40NTU	WP110404
Turbidity Calibration std, 1NTU	WP110408
Turbidity Calibration std, 5NTU	WP110407
Turbidity Calibration - CCV std, 10 NTU	WP110406
10 NTU Standard 500 ml	W3116

Intercept: 0.0585

Slope: 0.9594

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.145	0.09		10/25/2024	08:30
2	CAL2	1	1	1.120	1.11	10.6	10/25/2024	08:34
3	CAL3	5	1	4.600	4.73	-5.3	10/25/2024	08:38
4	CAL4	10	1	9.600	9.95	-0.5	10/25/2024	08:42
5	CAL5	20	1	19.400	20.16	0.8	10/25/2024	08:46
6	CAL6	40	1	38.400	39.96	-0.1	10/25/2024	08:50



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB133121

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.500	9.841	10/25/2024	09:02
2	ICB		1	0.108	0.052	10/25/2024	09:06
3	CCV1	10	1	9.120	9.445	10/25/2024	09:10
4	CCB1		1	0.115	0.059	10/25/2024	09:14
5	LB133121BL		1	0.100	0.043	10/25/2024	09:18
6	P4368-09		1	0.510	0.471	10/25/2024	09:22
7	CCV2	10	1	9.160	9.487	10/25/2024	09:46
8	CCB2		1	0.106	0.050	10/25/2024	09:50