



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/28/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:25
In Date: 10/25/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 10/26/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133127

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4484-01	MYE4H0	1	1.15	8.36	9.51	9.25	96.9	
P4484-02	MYE4H1	2	1.15	8.70	9.85	9.77	99.1	
P4484-03	MYE4H2	3	1.15	8.67	9.82	9.41	95.3	
P4484-04	MYE4H2D	4	1.15	8.67	9.82	9.41	95.3	
P4484-05	MYE4H2S	5	1.15	8.67	9.82	9.41	95.3	
P4484-06	MYE4H3	6	1.15	8.45	9.6	8.95	92.3	
P4484-07	MYE4H4	7	1.14	8.40	9.54	9.47	99.2	
P4484-08	MYE4H5	8	1.14	8.47	9.61	9.33	96.7	
P4484-09	MYE4H6	9	1.12	8.40	9.52	9.39	98.5	
P4484-10	MYE4H8	10	1.13	8.73	9.86	9.67	97.8	
P4484-11	MYE4H9	11	1.15	8.40	9.55	9.33	97.4	
P4484-12	MYE4J0	12	1.15	8.78	9.93	9.74	97.8	
P4484-13	MYE4J1	13	1.15	8.57	9.72	9.55	98.0	
P4484-14	MYE4J2	14	1.15	8.51	9.66	9.51	98.2	
P4484-15	MYE4J3	15	1.17	8.48	9.65	9.56	98.9	
P4484-16	MYE4J4	16	1.17	8.42	9.59	9.44	98.2	
P4484-17	MYE4J5	17	1.17	8.36	9.53	9.22	96.3	
P4484-18	MYE4J6	18	1.17	8.61	9.78	9.47	96.4	
P4484-19	MYE4J7	19	1.17	8.60	9.77	9.63	98.4	
P4484-20	MYE4J8	20	1.16	8.78	9.94	9.7	97.3	
P4484-21	MYE4J9	21	1.16	8.61	9.77	9.34	95.0	
P4484-22	MYE4L3	22	1.17	8.60	9.77	8.55	85.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$