



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/28/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:05
In Date: 10/25/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:36
Out Date: 10/26/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133128

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4519-01	MYE4K0	1	1.17	8.57	9.74	9.53	97.5	
P4519-02	MYE4K1	2	1.15	8.46	9.61	9.42	97.8	
P4519-03	MYE4K2	3	1.14	8.57	9.71	9.53	97.9	
P4519-04	MYE4K2D	4	1.14	8.57	9.71	9.53	97.9	
P4519-05	MYE4K2S	5	1.14	8.57	9.71	9.53	97.9	
P4519-06	MYE4K3	6	1.15	8.51	9.66	9.45	97.5	
P4519-07	MYE4K4	7	1.15	8.68	9.83	9.58	97.1	
P4519-08	MYE4K5	8	1.15	8.50	9.65	9.42	97.3	
P4519-09	MYE4K6	9	1.17	8.60	9.77	9.36	95.2	
P4519-10	MYE4K7	10	1.15	8.51	9.66	9.34	96.2	
P4519-11	MYE4K8	11	1.16	8.35	9.51	8.82	91.7	
P4519-12	MYE4K9	12	1.16	8.65	9.81	9.04	91.1	
P4519-13	MYE4L0	13	1.17	8.58	9.75	9.36	95.5	
P4519-14	MYE4L1	14	1.16	8.50	9.66	9.16	94.1	
P4519-15	MYE4L2	15	1.16	8.63	9.79	9.27	94.0	
P4519-16	MYE4P6	16	1.16	8.58	9.74	9.06	92.1	
P4519-17	MYE4P7	17	1.17	8.36	9.53	9.35	97.8	
P4519-18	MYE4P8	18	1.17	8.42	9.59	9.4	97.7	
P4519-19	MYE4P9	19	1.17	8.62	9.79	9.68	98.7	
P4519-20	MYE4Q0	20	1.16	8.67	9.83	9.44	95.5	
P4519-21	MYE4Q1	21	1.18	8.63	9.81	9.65	98.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$