

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/28/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 11:35
In Date: 10/26/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:15
Out Date: 10/27/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133140

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4494-01	A4F35	1	1.15	8.83	9.98	4.31	35.8	
P4494-02	A4F35MS	2	1.15	8.83	9.98	4.31	35.8	
P4494-03	A4F35MSD	3	1.15	8.83	9.98	4.31	35.8	
P4494-04	VHBLK001	4	1.00	1.00	2.00	2.00	100.0	vhblk
P4568-01	A4EA3	5	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P4568-02	A4EA4	6	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P4568-03	A4EF5	7	1.17	8.73	9.9	4.6	39.3	
P4568-04	A4EF5MS	8	1.17	8.73	9.9	4.6	39.3	
P4568-05	A4EF5MSD	9	1.17	8.73	9.9	4.6	39.3	
P4568-06	A4EF6	10	1.15	8.82	9.97	8.07	78.5	
P4568-07	A4EF8	11	1.18	8.80	9.98	4.87	41.9	
P4568-08	A4EF9	12	1.16	8.45	9.61	4.64	41.2	
P4568-09	A4EG0	13	1.16	8.81	9.97	5.78	52.4	
P4568-10	A4EH4	14	1.15	8.62	9.77	7.3	71.3	
P4568-11	A4EK1	15	1.15	8.83	9.98	6.69	62.7	
P4568-12	A4EE2	16	1.18	8.40	9.58	8.02	81.4	
P4568-13	A4EE3	17	1.15	8.78	9.93	6.27	58.3	
P4568-14	A4EE4	18	1.18	8.34	9.52	7.88	80.3	
P4568-15	A4EE5	19	1.17	8.81	9.98	8.94	88.2	
P4568-16	A4EE6	20	1.16	8.81	9.97	5.31	47.1	
P4568-17	A4EF1	21	1.18	8.60	9.78	7.2	70.0	
P4568-18	A4EF2	22	1.17	8.81	9.98	6.74	63.2	
P4568-19	A4EF3	23	1.18	8.79	9.97	6.61	61.8	
P4568-20	A4EF4	24	1.18	8.79	9.97	7.1	67.3	
P4568-21	A4EF7	25	1.17	8.79	9.96	7.29	69.6	
P4568-24	VHBLK002	26	1.00	1.00	2.00	2.00	100.0	vhblk

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$