



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/28/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:05
In Date: 10/26/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:22
Out Date: 10/27/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133141

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4520-01	MYE4L4	1	1.16	8.51	9.67	9.15	93.9	
P4520-02	MYE4L5	2	1.16	8.72	9.88	9.26	92.9	
P4520-03	MYE4L6	3	1.16	8.52	9.68	9.43	97.1	
P4520-04	MYE4L7	4	1.17	8.48	9.65	9.4	97.1	
P4520-05	MYE4L8	5	1.17	8.73	9.9	9.66	97.3	
P4520-06	MYE4L9	6	1.16	8.61	9.77	9.41	95.8	
P4520-07	MYE4M0	7	1.17	8.54	9.71	7.79	77.5	
P4520-08	MYE4M1	8	1.15	8.67	9.82	8.4	83.6	
P4520-09	MYE4M1D	9	1.15	8.67	9.82	8.4	83.6	
P4520-10	MYE4M1S	10	1.15	8.67	9.82	8.4	83.6	
P4520-11	MYE4M2	11	1.15	8.83	9.98	7.28	69.4	
P4520-12	MYE4M4	12	1.16	8.46	9.62	9.26	95.7	
P4520-13	MYE4M5	13	1.15	8.44	9.59	8.39	85.8	
P4520-14	MYE4M6	14	1.14	8.63	9.77	9.41	95.8	
P4520-15	MYE4M7	15	1.12	8.76	9.88	9.55	96.2	
P4520-16	MYE4M8	16	1.16	8.59	9.75	9.4	95.9	
P4520-17	MYE4M9	17	1.17	8.60	9.77	8.27	82.6	
P4520-18	MYE4N0	18	1.17	8.52	9.69	8.94	91.2	
P4520-19	MYE4N1	19	1.14	8.47	9.61	9.12	94.2	
P4520-20	MYE4N2	20	1.14	8.84	9.98	9.63	96.0	
P4520-21	MYE4N3	21	1.14	8.42	9.56	9.21	95.8	
P4520-22	MYE4N4	22	1.15	8.74	9.89	9.53	95.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$