



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/28/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:45
In Date: 10/26/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:28
Out Date: 10/27/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133142

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4521-01	MYE4N5	1	1.14	8.74	9.88	9.75	98.5	
P4521-02	MYE4N6	2	1.14	8.78	9.92	9.65	96.9	
P4521-03	MYE4N7	3	1.15	8.74	9.89	9.75	98.4	
P4521-04	MYE4N8	4	1.15	8.38	9.53	9.43	98.8	
P4521-05	MYE4N9	5	1.13	8.54	9.67	9.63	99.5	
P4521-06	MYE4P0	6	1.13	8.58	9.71	8.73	88.6	
P4521-07	MYE4P1	7	1.12	8.41	9.53	8.98	93.5	
P4521-08	MYE4P2	8	1.15	8.47	9.62	9.33	96.6	
P4521-09	MYE4P2D	9	1.15	8.47	9.62	9.33	96.6	
P4521-10	MYE4P2S	10	1.15	8.47	9.62	9.33	96.6	
P4521-11	MYE4P3	11	1.13	8.49	9.62	9.1	93.9	
P4521-12	MYE4P4	12	1.13	8.59	9.72	9.09	92.7	
P4521-13	MYE4P5	13	1.13	8.64	9.77	9.68	99.0	
P4521-14	MYE4Q2	14	1.13	8.40	9.53	9.33	97.6	
P4521-15	MYE4Q3	15	1.13	8.65	9.78	9.47	96.4	
P4521-16	MYE4Q4	16	1.13	8.71	9.84	9.48	95.9	
P4521-17	MYE4Q5	17	1.14	8.80	9.94	9.61	96.3	
P4521-18	MYE4Q6	18	1.15	8.73	9.88	9.69	97.8	
P4521-19	MYE4Q7	19	1.14	8.56	9.7	9.6	98.8	
P4521-20	MYE4Q8	20	1.15	8.53	9.68	9.46	97.4	
P4521-21	MYE4Q9	21	1.13	8.79	9.92	9.37	93.7	
P4521-22	MYE4R0	22	1.13	8.57	9.7	9.53	98.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$