



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/28/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:35
In Date: 10/26/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 10/27/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133143

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4522-01	MYE4R1	1	1.15	8.68	9.83	9.7	98.5	
P4522-02	MYE4R1D	2	1.15	8.68	9.83	9.7	98.5	
P4522-03	MYE4R1S	3	1.15	8.68	9.83	9.7	98.5	
P4522-04	MYE4R2	4	1.16	8.60	9.76	9.6	98.1	
P4522-05	MYE4R3	5	1.16	8.73	9.89	9.71	97.9	
P4522-06	MYE4R4	6	1.16	8.53	9.69	9.52	98.0	
P4522-07	MYE4R5	7	1.15	8.47	9.62	9.5	98.6	
P4522-08	MYE4R6	8	1.16	8.76	9.92	9.72	97.7	
P4522-09	MYE4R7	9	1.14	8.65	9.79	9.55	97.2	
P4522-10	MYE4R8	10	1.15	8.77	9.92	9.77	98.3	
P4522-11	MYE4S0	11	1.16	8.39	9.55	9.4	98.2	
P4522-12	MYE4S1	12	1.15	8.71	9.86	9.72	98.4	
P4522-13	MYE4S2	13	1.16	8.67	9.83	9.66	98.0	
P4522-14	MYE4S3	14	1.16	8.47	9.63	9.47	98.1	
P4522-15	MYE4S4	15	1.16	8.73	9.89	9.75	98.4	
P4522-16	MYE4S5	16	1.15	8.67	9.82	9.65	98.0	
P4522-17	MYE4S6	17	1.15	8.50	9.65	9.42	97.3	
P4522-18	MYE4S7	18	1.15	8.61	9.76	9.56	97.7	
P4522-19	MYE4S8	19	1.16	8.49	9.65	9.55	98.8	
P4522-20	MYE4S9	20	1.16	8.59	9.75	9.61	98.4	
P4522-21	MYE4T0	21	1.16	8.54	9.7	9.39	96.4	
P4522-22	MYE4T1	22	1.15	8.47	9.62	9.34	96.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$