



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/28/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 10/27/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 10/28/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133154

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4569-01	MYE516	1	1.13	8.45	9.58	9.21	95.6	
P4569-02	MYE517	2	1.18	8.45	9.63	9.28	95.9	
P4569-03	MYE518	3	1.17	8.56	9.73	9.39	96.0	
P4569-04	MYE519	4	1.12	8.80	9.92	9.78	98.4	
P4569-05	MYE519D	5	1.12	8.80	9.92	9.78	98.4	
P4569-06	MYE519S	6	1.12	8.80	9.92	9.78	98.4	
P4569-07	MYE520	7	1.15	8.38	9.53	9.26	96.8	
P4569-08	MYE521	8	1.17	8.60	9.77	9.57	97.7	
P4569-09	MYE522	9	1.18	8.66	9.84	9.31	93.9	
P4569-10	MYE523	10	1.14	8.79	9.93	9.75	98.0	
P4569-11	MYE524	11	1.17	8.58	9.75	9.52	97.3	
P4569-12	MYE525	12	1.17	8.41	9.58	9.37	97.5	
P4569-13	MYE526	13	1.16	8.56	9.72	9.43	96.6	
P4569-14	MYE527	14	1.18	8.53	9.71	9.28	95.0	
P4569-15	MYE528	15	1.16	8.40	9.56	9.14	95.0	
P4569-16	MYE529	16	1.18	8.42	9.6	9.11	94.2	
P4569-17	MYE530	17	1.14	8.45	9.59	9.33	96.9	
P4569-18	MYE532	18	1.12	8.70	9.82	9.57	97.1	
P4569-19	MYE533	19	1.18	8.52	9.7	9.24	94.6	
P4569-20	MYE534	20	1.18	8.48	9.66	9.24	95.0	
P4569-21	MYE535	21	1.18	8.63	9.81	9.59	97.5	
P4569-22	MYE536	22	1.17	8.80	9.97	9.73	97.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$