



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/28/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 10/27/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:49
Out Date: 10/28/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133156

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4571-01	MYE543	1	1.15	8.53	9.68	9.48	97.7	
P4571-02	MYE543D	2	1.15	8.53	9.68	9.48	97.7	
P4571-03	MYE543S	3	1.15	8.53	9.68	9.48	97.7	
P4571-04	MYE576	4	1.15	8.80	9.95	9.51	95.0	
P4571-05	MYE577	5	1.18	8.63	9.81	9.5	96.4	
P4571-06	MYE578	6	1.18	8.44	9.62	9.35	96.8	
P4571-07	MYE579	7	1.15	8.38	9.53	9.06	94.4	
P4571-08	MYE580	8	1.15	8.44	9.59	9.11	94.3	
P4571-09	MYE581	9	1.12	8.74	9.86	9.36	94.3	
P4571-10	MYE544	10	1.15	8.47	9.62	9.45	98.0	
P4571-11	MYE545	11	1.14	8.57	9.71	9.55	98.1	
P4571-12	MYE546	12	1.15	8.50	9.65	9.21	94.8	
P4571-13	MYE547	13	1.11	8.47	9.58	9.35	97.3	
P4571-14	MYE548	14	1.15	8.62	9.77	9.44	96.2	
P4571-15	MYE549	15	1.2	8.62	9.82	9.38	94.9	
P4571-16	MYE550	16	1.15	8.62	9.77	9.59	97.9	
P4571-17	MYE551	17	1.15	8.70	9.85	9.66	97.8	
P4571-18	MYE552	18	1.12	8.67	9.79	8.87	89.4	
P4571-19	MYE553	19	1.14	8.56	9.7	9.47	97.3	
P4571-20	MYE554	20	1.18	8.62	9.8	9.62	97.9	
P4571-21	MYE555	21	1.14	8.49	9.63	9.44	97.8	
P4571-22	MYE556	22	1.14	8.57	9.71	9.38	96.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$