



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/28/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:10
In Date: 10/27/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:55
Out Date: 10/28/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133157

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4572-01	MYE557	1	1.14	8.47	9.61	9.33	96.7	
P4572-02	MYE558	2	1.15	8.60	9.75	9.39	95.8	
P4572-03	MYE559	3	1.18	8.48	9.66	9.37	96.6	
P4572-04	MYE560	4	1.15	8.69	9.84	9.61	97.4	
P4572-05	MYE561	5	1.18	8.64	9.82	9.02	90.7	
P4572-06	MYE562	6	1.15	8.40	9.55	9.35	97.6	
P4572-07	MYE582	7	1.19	8.43	9.62	9.21	95.1	
P4572-08	MYE583	8	1.12	8.70	9.82	9.34	94.5	
P4572-09	MYE584	9	1.15	8.74	9.89	8.96	89.4	
P4572-10	MYE585	10	1.14	8.67	9.81	9.46	96.0	
P4572-11	MYE586	11	1.18	8.66	9.84	9.5	96.1	
P4572-12	MYE586D	12	1.18	8.66	9.84	9.5	96.1	
P4572-13	MYE586S	13	1.18	8.66	9.84	9.5	96.1	
P4572-14	MYE587	14	1.12	8.77	9.89	9.44	94.9	
P4572-15	MYE588	15	1.13	8.74	9.87	9.41	94.7	
P4572-16	MYE589	16	1.12	8.38	9.5	9.3	97.6	
P4572-17	MYE590	17	1.13	8.76	9.89	9.47	95.2	
P4572-18	MYE591	18	1.18	8.40	9.58	9.23	95.8	
P4572-19	MYE592	19	1.15	8.59	9.74	9.4	96.0	
P4572-20	MYE593	20	1.12	8.72	9.84	9.29	93.7	
P4572-21	MYE594	21	1.18	8.47	9.65	9.27	95.5	
P4572-22	MYE595	22	1.12	8.71	9.83	9.46	95.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$