

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133177

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/28/2024	16:10
2	CAL2	1	Water	NA	NA	20.2	7.00	10/28/2024	16:11
3	CAL3	1	Water	NA	NA	20.3	10.02	10/28/2024	16:12
4	ICV	1	Water	NA	NA	20.2	7.00	10/28/2024	16:15
5	CCV1	1	Water	NA	NA	20.2	2.01	10/28/2024	16:20
6	P4596-01	1	Water	NA	NA	20.1	8.44	10/28/2024	16:25
7	P4596-01DUP	1	Water	NA	NA	20.2	8.45	10/28/2024	16:26
8	P4596-02	1	Water	NA	NA	20.3	7.80	10/28/2024	16:30
9	P4596-03	1	Water	NA	NA	20.7	7.86	10/28/2024	16:37
10	P4596-05	1	Water	NA	NA	20.2	9.02	10/28/2024	16:45
11	P4596-06	1	Water	NA	NA	20.2	8.45	10/28/2024	16:50
12	CCV2	1	Water	NA	NA	20.3	12.02	10/28/2024	16:55

WORKLIST(Hardcopy Internal Chain)

133177

WorkList Name : ph water p4586 WorkList ID : 184882 Department : Wet-Chemistry Date : 10-28-2024 15:53:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4596-01	50445	Water	pH	Cool 4 deg C	PSEG03	K61	10/28/2024	9040C
P4596-02	50446	Water	pH	Cool 4 deg C	PSEG03	K61	10/28/2024	9040C
P4596-03	50447	Water	pH	Cool 4 deg C	PSEG03	K61	10/28/2024	9040C
P4596-05	TOTE-WASTE-WATER-50468	Water	pH	Cool 4 deg C	PSEG03	K61	10/28/2024	9040C
P4596-06	TOTE-WASTE-WATER-50469	Water	pH	Cool 4 deg C	PSEG03	K61	10/28/2024	9040C

Date/Time 10/28/24 16:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 10/28/24 18:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]