



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/30/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:55
In Date: 10/29/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/30/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133183

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4588-01	MYE5J2	1	1.15	8.41	9.56	7.4	74.3	
P4588-02	MYE5J3	2	1.15	8.79	9.94	9.41	94.0	
P4588-03	MYE5J4	3	1.15	8.68	9.83	8.54	85.1	
P4588-04	MYE5J5	4	1.14	8.51	9.65	9.43	97.4	
P4588-05	MYE5J6	5	1.16	8.72	9.88	8.37	82.7	
P4588-06	MYE5J7	6	1.14	8.59	9.73	8.52	85.9	
P4588-07	MYE5J8	7	1.15	8.63	9.78	9.48	96.5	
P4588-08	MYE5J9	8	1.13	8.46	9.59	9.00	93.0	
P4588-09	MYE5K0	9	1.12	8.55	9.67	8.74	89.1	
P4588-10	MYE5K1	10	1.12	8.59	9.71	8.71	88.4	
P4588-11	MYE5K2	11	1.13	8.75	9.88	9.4	94.5	
P4588-12	MYE5K3	12	1.13	8.39	9.52	9.1	95.0	
P4588-13	MYE5K4	13	1.13	8.69	9.82	8.87	89.1	
P4588-14	MYE5K5	14	1.12	8.42	9.54	9.07	94.4	
P4588-15	MYE5K5D	15	1.12	8.42	9.54	9.07	94.4	
P4588-16	MYE5K5S	16	1.12	8.42	9.54	9.07	94.4	
P4588-17	MYE5K6	17	1.13	8.67	9.8	9.63	98.0	
P4588-18	MYE5K7	18	1.12	8.56	9.68	9.27	95.2	
P4588-19	MYE5K8	19	1.13	8.85	9.98	7.88	76.3	
P4588-20	MYE5K9	20	1.14	8.83	9.97	9.47	94.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$