


LB133185

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.1900e-004

C= 2.0230e-002

Rho= 0.9987532

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.023	0.023	11	0.000	11					1
0.2	0.200	0.215	0.015	889	0.0 %	889					2
2.5	2.500	2.285	-0.215	10342	0.0 %	10342					3
5.0	5.000	5.154	0.154	23440	0.0 %	23440					4
7.5	7.500	7.769	0.269	35384	0.0 %	35384					5
10.0	10.000	9.754	-0.246	44446	0.0 %	44446					2

LB133185 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Oct 2024 10:07:32

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	11	-	0.0000	7471B	PPB	29 Oct 2024 10:10:33	S	Std
0.2 - 1	50.2	889	-	0.2000	7471B	PPB	29 Oct 2024 10:12:51	S	Std
2.5 - 1	52.5	10342	-	2.5000	7471B	PPB	29 Oct 2024 10:15:07	S	Std
5.0 - 1	55	23440	-	5.0000	7471B	PPB	29 Oct 2024 10:17:24	S	Std
7.5 - 1	57.5	35384	-	7.5000	7471B	PPB	29 Oct 2024 10:19:42	S	Std
10.0 - 1	57.0	44446	-	10.0000	7471B	PPB	29 Oct 2024 10:29:33	S	Std
ICV57 - 1	ICV57	18909	4.1614	-	7471B	PPB	29 Oct 2024 10:32:20	U	SMPL
ICB57 - 1	ICB57	-370	-0.0608	-	7471B	PPB	29 Oct 2024 10:34:36	U	SMPL
CCV86 - 1	CCV86	23331	5.1298	-	7471B	PPB	29 Oct 2024 10:36:54	U	SMPL
CCB86 - 1	CCB86	-414	-0.0704	-	7471B	PPB	29 Oct 2024 10:39:08	U	SMPL
CRA - 1	CRA	805	0.1965	-	7471B	PPB	29 Oct 2024 10:41:26	U	SMPL
HighStd - 1	HighStd	44910	9.8557	-	7471B	PPB	29 Oct 2024 10:43:40	U	SMPL
ChkStd - 1	ChkStd	32476	7.1326	-	7471B	PPB	29 Oct 2024 10:45:55	U	SMPL
PB164487BL - 1	PBS	-427	-0.0733	-	7471B	PPB	29 Oct 2024 10:48:16	U	SMPL
PB164487BS - 1	LCSS	15933	3.5096	-	7471B	PPB	29 Oct 2024 10:50:37	U	SMPL
P4574-01 - 1	GRAVEL-1	-55	0.0082	-	7471B	PPB	29 Oct 2024 10:52:53	U	SMPL
P4574-01DUP - 1	GRAVEL-1DUP	81	0.0380	-	7471B	PPB	29 Oct 2024 10:55:14	U	SMPL
P4574-01MS - 1	GRAVEL-1MS	18635	4.0795	-	7471B	PPB	29 Oct 2024 10:57:31	U	SMPL
P4574-01MSD - 1	GRAVEL-1MSD	18788	4.1349	-	7471B	PPB	29 Oct 2024 10:59:48	U	SMPL
P4574-04 - 1	GRAVEL-2	368	0.1008	-	7471B	PPB	29 Oct 2024 11:02:05	U	SMPL
CCV87 - 1	CCV87	23549	5.1776	-	7471B	PPB	29 Oct 2024 11:04:23	U	SMPL
CCB87 - 1	CCB87	-396	-0.0665	-	7471B	PPB	29 Oct 2024 11:06:38	U	SMPL
P4574-07 - 1	CONCRETE-1	59	0.0332	-	7471B	PPB	29 Oct 2024 11:08:56	U	SMPL
P4574-09 - 1	CONCRETE-2	-11	0.0178	-	7471B	PPB	29 Oct 2024 11:11:12	U	SMPL
P4574-11 - 1	CONCRETE-3	197	0.0634	-	7471B	PPB	29 Oct 2024 11:13:27	U	SMPL
P4575-01 - 1	PL-02-102424	9114	2.0162	-	7471B	PPB	29 Oct 2024 11:15:42	U	SMPL
P4577-01 - 1	TR-05-102524	4314	0.9650	-	7471B	PPB	29 Oct 2024 11:17:58	U	SMPL
P4578-03 - 1	WB-305-TOP	169263	37.0896	-	7471B	PPB	29 Oct 2024 11:20:17	U	SMPL
P4578-04 - 1	WB-305-TOP-1	133265	29.2058	-	7471B	PPB	29 Oct 2024 11:22:36	U	SMPL
P4578-05 - 1	WB-305-BOT	4440	0.9926	-	7471B	PPB	29 Oct 2024 11:25:15	U	SMPL
P4578-07 - 1	WB-305-BOT-1	2202	0.5025	-	7471B	PPB	29 Oct 2024 11:27:53	U	SMPL
P4594-01 - 1	TP-4	3	0.0209	-	7471B	PPB	29 Oct 2024 11:30:10	U	SMPL
CCV88 - 1	CCV88	24018	5.2803	-	7471B	PPB	29 Oct 2024 11:32:27	U	SMPL
CCB88 - 1	CCB88	-179	-0.0190	-	7471B	PPB	29 Oct 2024 11:34:43	U	SMPL
P4594-05 - 1	BP-F17	591	0.1497	-	7471B	PPB	29 Oct 2024 11:37:01	U	SMPL
P4594-09 - 1	BP-F16	594	0.1503	-	7471B	PPB	29 Oct 2024 11:39:16	U	SMPL
P4594-13 - 1	TP-5	23	0.0253	-	7471B	PPB	29 Oct 2024 11:41:31	U	SMPL
P4594-17 - 1	BP-F15	213	0.0669	-	7471B	PPB	29 Oct 2024 11:43:47	U	SMPL
P4595-01 - 1	PITKIN-COMP	12028	2.6544	-	7471B	PPB	29 Oct 2024 11:46:02	U	SMPL
P4574-01LX5 - 1		-281	-0.0413	-	7471B	PPB	29 Oct 2024 11:48:18	U	SMPL
P4574-01A - 1		17102	3.7656	-	7471B	PPB	29 Oct 2024 11:50:37	U	SMPL
P4578-03DLX10 - 1		17872	3.9343	-	7471B	PPB	29 Oct 2024 11:52:53	U	SMPL
P4578-04DLX10 - 1		11625	2.5662	-	7471B	PPB	29 Oct 2024 11:55:13	U	SMPL
CCV89 - 1	CCV89	23913	5.2573	-	7471B	PPB	29 Oct 2024 11:57:31	U	SMPL
CCB89 - 1	CCB89	-261	-0.0369	-	7471B	PPB	29 Oct 2024 11:59:48	U	SMPL