

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133189

Ambient Barometric Pressure (mmHg): 781.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.1	10/29/2024	11:00
2	P4546-03		1	11	100.00	>212.0	>212.0	10/29/2024	11:30
3	P4546-03DUP		1	11	100.00	>212.0	>212.0	10/29/2024	12:00
4	P4564-01		1	14	100.00	>212.0	>212.0	10/29/2024	12:30
5	P4596-04		1	12	100.00	>212.0	>212.0	10/29/2024	13:00
6	P4596-07		1	13	100.00	>212.0	>212.0	10/29/2024	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66133189

WorkList Name : FP1028

WorkList ID : 184865

Department : Wet-Chemistry

Date : 10-28-2024 11:09:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4546-03	34839-40	Water	Flash Point	Cool 4 deg C	PSEG03	K61	10/24/2024	1010B
P4564-01	100324	Water	Flash Point	Cool 4 deg C	PSEG03	K63	10/25/2024	1010B
P4596-04	DL-1	Water	Flash Point	Cool 4 deg C	PSEG03	K61	10/28/2024	1010B
P4596-07	TOTE-WATER-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	K61	10/28/2024	1010B

Date/Time 10/29/2024 10:45
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JH(DC)

Date/Time 10/29/2024 14:10
 Raw Sample Received by: JH(DC)
 Raw Sample Relinquished by: RM CWC