

LB133205


7471B

INSTRUMENT ID: CY1

Linear

μ Abs.:

47073



A= 0.0000e+000

B= 2.1129e-004

C= -9.9334e-003

Rho= 0.9999112

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.005	0.005	72	0.000	72					
0.2	0.200	0.199	-0.001	991	0.0 %	991					
2.5	2.500	2.465	-0.035	11712	0.0 %	11712					
5.0	5.000	4.998	-0.002	23703	0.0 %	23703					
7.5	7.500	7.596	0.096	35996	0.0 %	35996					
10.0	10.000	9.936	-0.064	47073	0.0 %	47073					

LB133205 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Oct 2024 11:34:50

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	72	-	0.0000	7471B	PPB	30 Oct 2024 11:52:57	S	Std
0.2 - 1	50.2	991	-	0.2000	7471B	PPB	30 Oct 2024 11:55:14	S	Std
2.5 - 1	52.5	11712	-	2.5000	7471B	PPB	30 Oct 2024 11:57:31	S	Std
5.0 - 1	55	23703	-	5.0000	7471B	PPB	30 Oct 2024 11:59:48	S	Std
7.5 - 1	57.5	35996	-	7.5000	7471B	PPB	30 Oct 2024 12:02:06	S	Std
10.0 - 1	510	47073	-	10.0000	7471B	PPB	30 Oct 2024 12:04:25	S	Std
ICV58 - 1	ICV58	19594	4.1302	-	7471B	PPB	30 Oct 2024 12:07:53	U	SMPL
ICB58 - 1	ICB58	-291	-0.0714	-	7471B	PPB	30 Oct 2024 12:10:08	U	SMPL
CCV90 - 1	CCV90	23948	5.0501	-	7471B	PPB	30 Oct 2024 12:12:25	U	SMPL
CCB90 - 1	CCB90	-302	-0.0737	-	7471B	PPB	30 Oct 2024 12:14:40	U	SMPL
CRA - 1	CRA	877	0.1754	-	7471B	PPB	30 Oct 2024 12:16:57	U	SMPL
HighStd - 1	HighStd	46733	9.8645	-	7471B	PPB	30 Oct 2024 12:19:12	U	SMPL
ChkStd - 1	ChkStd	33346	7.0359	-	7471B	PPB	30 Oct 2024 12:21:27	U	SMPL
PB164521BL - 1	PBS	-77	-0.0262	-	7471B	PPB	30 Oct 2024 12:26:29	U	SMPL
PB164521BS - 1	LCSS	17345	3.6550	-	7471B	PPB	30 Oct 2024 12:28:44	U	SMPL
P4597-01 - 1	RED-1-1	3490	0.7275	-	7471B	PPB	30 Oct 2024 12:36:04	U	SMPL
P4597-01DUP - 1	RED-1-1DUP	3113	0.6478	-	7471B	PPB	30 Oct 2024 12:38:20	U	SMPL
P4597-01MS - 1	RED-1-1MS	22461	4.7359	-	7471B	PPB	30 Oct 2024 12:40:38	U	SMPL
P4597-01MSD - 1	RED-1-1MSD	20982	4.4234	-	7471B	PPB	30 Oct 2024 12:42:55	U	SMPL
P4597-04 - 1	RED-1-2	241	0.0410	-	7471B	PPB	30 Oct 2024 12:45:13	U	SMPL
CCV91 - 1	CCV91	24177	5.0985	-	7471B	PPB	30 Oct 2024 12:47:30	U	SMPL
CCB91 - 1	CCB91	-257	-0.0642	-	7471B	PPB	30 Oct 2024 12:49:45	U	SMPL
P4597-07 - 1	BLUE-2-1	2413	0.4999	-	7471B	PPB	30 Oct 2024 12:52:03	U	SMPL
P4597-10 - 1	BLUE-2-2	174	0.0268	-	7471B	PPB	30 Oct 2024 12:54:18	U	SMPL
P4598-01 - 1	BP-F12	176	0.0273	-	7471B	PPB	30 Oct 2024 12:56:34	U	SMPL
P4598-05 - 1	BP-F11	354	0.0649	-	7471B	PPB	30 Oct 2024 12:58:50	U	SMPL
P4598-09 - 1	TP-8	3814	0.7959	-	7471B	PPB	30 Oct 2024 13:01:06	U	SMPL
P4611-01 - 1	TP-1	43634	9.2097	-	7471B	PPB	30 Oct 2024 13:03:22	U	SMPL
P4611-04 - 1	TP-2	160631	33.9305	-	7471B	PPB	30 Oct 2024 13:05:39	U	SMPL
P4611-07 - 1	TP-3	66480	14.0369	-	7471B	PPB	30 Oct 2024 13:08:01	U	SMPL
P4611-10 - 1	TP-4	5131	1.0742	-	7471B	PPB	30 Oct 2024 13:10:35	U	SMPL
P4611-13 - 1	TP-5	5398	1.1306	-	7471B	PPB	30 Oct 2024 13:12:59	U	SMPL
CCV92 - 1	CCV92	24561	5.1797	-	7471B	PPB	30 Oct 2024 13:15:15	U	SMPL
CCB92 - 1	CCB92	-270	-0.0670	-	7471B	PPB	30 Oct 2024 13:17:32	U	SMPL
P4611-16 - 1	TP-6	130141	27.4881	-	7471B	PPB	30 Oct 2024 13:19:51	U	SMPL
P4597-01LX5 - 1		151	0.0220	-	7471B	PPB	30 Oct 2024 13:22:06	U	SMPL
P4597-01A - 1		21222	4.4742	-	7471B	PPB	30 Oct 2024 13:24:38	U	SMPL
P4611-04DLX10 - 1		18254	3.8470	-	7471B	PPB	30 Oct 2024 13:29:58	U	SMPL
P4611-07DLX5 - 1		12913	2.7185	-	7471B	PPB	30 Oct 2024 13:32:14	U	SMPL
P4611-16DLX10 - 1		13954	2.9385	-	7471B	PPB	30 Oct 2024 13:34:33	U	SMPL
CCV93 - 1	CCV93	24379	5.1412	-	7471B	PPB	30 Oct 2024 13:36:51	U	SMPL
CCB93 - 1	CCB93	-244	-0.0615	-	7471B	PPB	30 Oct 2024 13:39:10	U	SMPL