

LB133211

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

10/30/2024 14:25

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.004	0.0	0.112	
ICB1	-0.009	0.0	0.011	
CCV1	0.987	0.0	0.110	
CCB1	-0.013	0.0	0.011	
PB164503BL	-0.006	0.0	0.011	
PB164503BS	0.998	0.0	0.111	
P4368-12	0.026	0.0	0.014	
CCV2	0.997	0.0	0.111	
CCB2	-0.006	0.0	0.011	

N	9
Mean	0.442
SD	0.5262
CV%	119.06

Aquakem v. 7.2AQ1

Results from time period:

Wed Oct 30 14:17:05 2024

Wed Oct 30 14:23:00 2024

Sample Id	Sam/Ctr/cA	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0092	mg/l	10/30/2024 12:30:06	
0.05PPM	A	Phenolics-	P	0.0497	mg/l	10/30/2024 12:30:07	
0.1PPM	A	Phenolics-	P	0.0996	mg/l	10/30/2024 12:30:08	
0.25PPM	A	Phenolics-	P	0.256	mg/l	10/30/2024 12:30:09	
0.50PPM	A	Phenolics-	P	0.5035	mg/l	10/30/2024 12:30:10	
1.0PPM	A	Phenolics-	P	1.004	mg/l	10/30/2024 12:30:11	
2.0PPM	A	Phenolics-	P	1.9964	mg/l	10/30/2024 12:30:12	
ICV1	S	Phenolics-	P	1.0036	mg/l	10/30/2024 14:17:05	
ICB1	S	Phenolics-	P	-0.0088	mg/l	10/30/2024 14:17:07	
CCV1	S	Phenolics-	P	0.9873	mg/l	10/30/2024 14:17:10	
CCB1	S	Phenolics-	P	-0.013	mg/l	10/30/2024 14:17:11	
PB164503BL	S	Phenolics-	P	-0.0062	mg/l	10/30/2024 14:17:14	
PB164503BS	S	Phenolics-	P	0.9982	mg/l	10/30/2024 14:17:15	
P4368-12	S	Phenolics-	P	0.0258	mg/l	10/30/2024 14:22:56	
CCV2	S	Phenolics-	P	0.997	mg/l	10/30/2024 14:22:57	
CCB2	S	Phenolics-	P	-0.0061	mg/l	10/30/2024 14:23:00	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

10/30/2024 12:55

Test Phenolics-


Accepted 10/30/2024 12:34

Factor 10.05

Bias 0.012

Coeff. of det. 0.999949

Errors



	Calibrator	Response	Calc. con.	Conc.	Re Errors
1	0.00PPM	0.011	-0.0092	0.0000	-
2	Phenol-2	0.017	0.0497	0.0500	-0.6
3	Phenol-2	0.022	0.0996	0.1000	-0.4
4	Phenol-2	0.037	0.2560	0.2500	2.4
5	Phenol-2	0.062	0.5035	0.5000	0.7
6	Phenol-2	0.112	1.0040	1.0000	0.4
7	Phenol-2	0.210	1.9964	2.0000	-0.2

10/30/2024

RM