



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/31/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:50
In Date: 10/30/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 10/31/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133213

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4581-01	A4EC9	1	1.15	8.41	9.56	8.04	81.9	
P4581-02	A4EC9MS	2	1.15	8.41	9.56	8.04	81.9	
P4581-03	A4EC9MSD	3	1.15	8.41	9.56	8.04	81.9	
P4581-04	A4ED0	4	1.15	8.82	9.97	8.75	86.2	
P4581-05	A4ED1	5	1.12	8.77	9.89	8.52	84.4	
P4581-06	A4ED2	6	1.17	8.74	9.91	8.53	84.2	
P4581-07	VHBLK001	7	1.00	1.00	2.00	2.00	100.0	vhblk
P4581-08	A4EA8	8	1.14	8.83	9.97	8.4	82.2	
P4581-09	A4EC2	9	1.12	8.77	9.89	8.5	84.2	
P4581-10	A4EC4	10	1.16	8.58	9.74	8.03	80.1	
P4581-11	A4EC6	11	1.18	8.49	9.67	8.37	84.7	
P4581-12	A4EC7	12	1.15	8.42	9.57	8.29	84.8	
P4581-13	A4ED3	13	1.15	8.83	9.98	7.99	77.5	
P4581-14	A4ED5	14	1.18	8.79	9.97	8.62	84.6	
P4581-15	A4ED6	15	1.11	8.88	9.99	9.4	93.4	
P4581-16	A4ED7	16	1.14	8.82	9.96	8.2	80.0	
P4581-19	A4EA9	17	1.15	8.80	9.95	8.5	83.5	
P4581-20	A4EB1	18	1.12	8.56	9.68	8.7	88.6	
P4581-21	A4EB2	19	1.16	8.48	9.64	8.11	82.0	
P4581-22	A4EB4	20	1.19	8.55	9.74	8.24	82.5	
P4581-23	A4EB5	21	1.13	8.70	9.83	8.36	83.1	
P4581-24	A4EB7	22	1.18	8.79	9.97	8.77	86.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$