

7470A

LB138216

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34067

A= 0.0000e+000

B= 2.9557e-004

C= -1.5768e-004

Rho= 0.9998859

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.003	-0.003	-8	0.000	-8	0				-
0.2	0.200	0.198	-0.002	671	0.0 %	671					-1
2.5	2.500	2.545	0.045	8611	0.0 %	8611					2
5.0	5.000	4.999	-0.001	16915	0.0 %	16915					0
7.5	7.500	7.391	-0.109	25007	0.0 %	25007					-1
10.0	10.000	10.069	0.069	34067	0.0 %	34067					1

LB133216 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Oct 2024 15:29:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-8	-	0.0000	7470A	PPB	30 Oct 2024 15:33:15	S	Std
0.2 - 1	50.2	671	-	0.2000	7470A	PPB	30 Oct 2024 15:35:32	S	Std
2.5 - 1	52.5	8611	-	2.5000	7470A	PPB	30 Oct 2024 15:37:49	S	Std
5.0 - 1	55	16915	-	5.0000	7470A	PPB	30 Oct 2024 15:40:06	S	Std
7.5 - 1	59.5	25007	-	7.5000	7470A	PPB	30 Oct 2024 15:42:22	S	Std
10.0 - 1	50	34067	-	10.0000	7470A	PPB	30 Oct 2024 15:44:39	S	Std
ICV59 - 1	ICV59	12627	3.7320	-7470A	PPB	PPB	30 Oct 2024 15:47:54	U	SMPL
ICB59 - 1	ICB59	-242	-0.0717	-7470A	PPB	PPB	30 Oct 2024 15:50:09	U	SMPL
CCV94 - 1	CCV94	17066	5.0440	-7470A	PPB	PPB	30 Oct 2024 15:52:25	U	SMPL
CCB94 - 1	CCB94	-253	-0.0749	-7470A	PPB	PPB	30 Oct 2024 15:54:40	U	SMPL
CRA - 1	CRA	493	0.1456	-7470A	PPB	PPB	30 Oct 2024 15:56:57	U	SMPL
HighStd - 1	HighStd	33832	9.9995	-7470A	PPB	PPB	30 Oct 2024 15:59:11	U	SMPL
ChkStd - 1	ChkStd	22158	6.5490	-7470A	PPB	PPB	30 Oct 2024 16:01:26	U	SMPL
PB164520BL - 1	PBW	-212	-0.0628	-7470A	PPB	PPB	30 Oct 2024 16:06:34	U	SMPL
PB164520BS - 1	LCSW	14307	4.2285	-7470A	PPB	PPB	30 Oct 2024 16:08:49	U	SMPL
P4594-04 - 1	TP-4	-265	-0.0785	-7470A	PPB	PPB	30 Oct 2024 16:11:05	U	SMPL
P4594-08 - 1	BP-F17	-85	-0.0253	-7470A	PPB	PPB	30 Oct 2024 16:13:24	U	SMPL
P4594-12 - 1	BP-F16	-70	-0.0208	-7470A	PPB	PPB	30 Oct 2024 16:15:41	U	SMPL
P4594-16 - 1	TP-5	-38	-0.0114	-7470A	PPB	PPB	30 Oct 2024 16:17:59	U	SMPL
P4594-20 - 1	BP-F15	-45	-0.0135	-7470A	PPB	PPB	30 Oct 2024 16:20:15	U	SMPL
CCV95 - 1	CCV95	17003	5.0254	-7470A	PPB	PPB	30 Oct 2024 16:22:30	U	SMPL
CCB95 - 1	CCB95	-252	-0.0746	-7470A	PPB	PPB	30 Oct 2024 16:24:46	U	SMPL
P4595-02 - 1	PITKIN-COMP	-77	-0.0229	-7470A	PPB	PPB	30 Oct 2024 16:27:03	U	SMPL
P4597-03 - 1	RED-1-1	-65	-0.0194	-7470A	PPB	PPB	30 Oct 2024 16:29:18	U	SMPL
P4597-06 - 1	RED-1-2	-47	-0.0140	-7470A	PPB	PPB	30 Oct 2024 16:31:32	U	SMPL
P4597-09 - 1	BLUE-2-1	-71	-0.0211	-7470A	PPB	PPB	30 Oct 2024 16:33:48	U	SMPL
P4597-12 - 1	BLUE-2-2	-16	-0.0049	-7470A	PPB	PPB	30 Oct 2024 16:36:04	U	SMPL
P4598-04 - 1	BP-F12	-23	-0.0070	-7470A	PPB	PPB	30 Oct 2024 16:38:20	U	SMPL
P4598-08 - 1	BP-F11	-28	-0.0084	-7470A	PPB	PPB	30 Oct 2024 16:40:36	U	SMPL
P4598-12 - 1	TP-8	-39	-0.0117	-7470A	PPB	PPB	30 Oct 2024 16:42:53	U	SMPL
P4598-12DUP - 1	TP-8DUP	-21	-0.0064	-7470A	PPB	PPB	30 Oct 2024 16:54:37	U	SMPL
P4598-12MS - 1	TP-8MS	12195	3.6043	-7470A	PPB	PPB	30 Oct 2024 16:56:52	U	SMPL
CCV96 - 1	CCV96	17312	5.1167	-7470A	PPB	PPB	30 Oct 2024 16:59:07	U	SMPL
CCB96 - 1	CCB96	-286	-0.0847	-7470A	PPB	PPB	30 Oct 2024 17:01:24	U	SMPL
P4598-12MSD - 1	TP-8MSD	13099	3.8715	-7470A	PPB	PPB	30 Oct 2024 17:03:41	U	SMPL
PB164478TB - 1	PB164478TB	-241	-0.0714	-7470A	PPB	PPB	30 Oct 2024 17:05:56	U	SMPL
P4598-12LX5 - 1		-51	-0.0152	-7470A	PPB	PPB	30 Oct 2024 17:08:13	U	SMPL
P4598-12A - 1		13272	3.9226	-7470A	PPB	PPB	30 Oct 2024 17:10:29	U	SMPL
CCV97 - 1	CCV97	16741	4.9479	-7470A	PPB	PPB	30 Oct 2024 17:12:44	U	SMPL
CCB97 - 1	CCB97	-305	-0.0903	-7470A	PPB	PPB	30 Oct 2024 17:15:02	U	SMPL