


LB133224

SFAM01.1

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.5195e-004 slope

C= 2.7095e-002 $y = mx + c$

Rho= 0.9998322

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/10
0.00	0.000	0.032	0.032	20	0.000	20					17
0.05	0.050					442	0				17
0.20	0.200	0.233	0.033	816	0.0 %	816					0
2.50	2.500	2.497	-0.003	9802	0.0 %	9802					1
5.00	5.000	4.931	-0.069	19463	0.0 %	19463					1
7.50	7.500	7.403	-0.097	29274	0.0 %	29274					1
10.0	10.000	10.105	0.105	40000	0.0 %	40000					1

LB133224
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	20	-		0 SFAM01.1	PPB	10/30/2024 18:18	Std
	0.2 S01	816	-		0.2 SFAM01.1	PPB	10/30/2024 18:21	Std
	2.5 S02	9802	-		2.5 SFAM01.1	PPB	10/30/2024 18:23	Std
	5 S03	19463	-		5 SFAM01.1	PPB	10/30/2024 18:25	Std
	7.5 S04	29274	-		7.5 SFAM01.1	PPB	10/30/2024 18:31	Std
	10 S05	40000	-		10 SFAM01.1	PPB	10/30/2024 18:33	Std
ICV060	ICV060	16905	4.2863 -		SFAM01.1	PPB	10/30/2024 18:39	SMPL
ICB060	ICB060	-217	-0.0276 -		SFAM01.1	PPB	10/30/2024 18:41	SMPL
CCV098	CCV098	19593	4.9635 -		SFAM01.1	PPB	10/30/2024 18:44	SMPL
CCB098	CCB098	-220	-0.0283 -		SFAM01.1	PPB	10/30/2024 18:46	SMPL
PB164549BL	PBS549	59	0.042 -		SFAM01.1	PPB	10/30/2024 18:48	SMPL
P4499-01	MBHCJ7	1319	0.3594 -		SFAM01.1	PPB	10/30/2024 18:50	SMPL
P4499-02	MBHCJ8	392	0.1259 -		SFAM01.1	PPB	10/30/2024 18:53	SMPL
P4499-03	MBHCJ9	1497	0.4043 -		SFAM01.1	PPB	10/30/2024 18:55	SMPL
P4499-04	MBHCK0	1984	0.527 -		SFAM01.1	PPB	10/30/2024 18:57	SMPL
P4499-05	MBHCK1	880	0.2488 -		SFAM01.1	PPB	10/30/2024 18:59	SMPL
P4499-06	MBHCK2	672	0.1964 -		SFAM01.1	PPB	10/30/2024 19:02	SMPL
P4499-07	MBHCK8	784	0.2246 -		SFAM01.1	PPB	10/30/2024 19:04	SMPL
P4499-08	MBHCK9	2168	0.5733 -		SFAM01.1	PPB	10/30/2024 19:06	SMPL
P4499-09	MBHCL0	2392	0.6298 -		SFAM01.1	PPB	10/30/2024 19:08	SMPL
P4499-10	MBHCL1	2816	0.7366 -		SFAM01.1	PPB	10/30/2024 19:11	SMPL
P4499-11	MBHCL2	1387	0.3765 -		SFAM01.1	PPB	10/30/2024 19:13	SMPL
P4499-12	MBHCY0	574	0.1717 -		SFAM01.1	PPB	10/30/2024 19:15	SMPL
P4499-13	MBHCL8	358	0.1173 -		SFAM01.1	PPB	10/30/2024 19:18	SMPL
P4499-14	MBHCL9	598	0.1778 -		SFAM01.1	PPB	10/30/2024 19:20	SMPL
P4499-15	MBHCM0	1191	0.3272 -		SFAM01.1	PPB	10/30/2024 19:22	SMPL
P4499-16	MBHCM1	619	0.1831 -		SFAM01.1	PPB	10/30/2024 19:24	SMPL
P4499-17	MBHCM2	801	0.2289 -		SFAM01.1	PPB	10/30/2024 19:27	SMPL
P4499-18	MBHCM8	714	0.207 -		SFAM01.1	PPB	10/30/2024 19:29	SMPL
P4499-19	MBHCM9	635	0.1871 -		SFAM01.1	PPB	10/30/2024 19:31	SMPL
CCV099	CCV099	20240	5.1265 -		SFAM01.1	PPB	10/30/2024 19:33	SMPL
CCB099	CCB099	-79	0.0072 -		SFAM01.1	PPB	10/30/2024 19:36	SMPL
P4499-20	MBHCN0	1029	0.2864 -		SFAM01.1	PPB	10/30/2024 19:38	SMPL
P4499-21	MBHCN0D	814	0.2322 -		SFAM01.1	PPB	10/30/2024 19:40	SMPL
P4499-22	MBHCN0S	10838	2.7577 -		SFAM01.1	PPB	10/30/2024 19:42	SMPL
PB164550BL	PBS551	-87	0.0052 -		SFAM01.1	PPB	10/30/2024 19:45	SMPL
P4497-01	MBHCY5	1444	0.3909 -		SFAM01.1	PPB	10/30/2024 19:47	SMPL
P4497-02	MBHCZ5	937	0.2632 -		SFAM01.1	PPB	10/30/2024 19:49	SMPL
P4497-03	MBHDA3	858	0.2433 -		SFAM01.1	PPB	10/30/2024 19:52	SMPL
P4497-04	MBHDA4	1383	0.3755 -		SFAM01.1	PPB	10/30/2024 19:54	SMPL
P4497-05	MBHDA5	495	0.1518 -		SFAM01.1	PPB	10/30/2024 19:56	SMPL
P4497-06	MBHDA6	733	0.2118 -		SFAM01.1	PPB	10/30/2024 19:58	SMPL
P4497-07	MBHDA7	928	0.2609 -		SFAM01.1	PPB	10/30/2024 20:01	SMPL
P4497-08	MBHDB3	648	0.1904 -		SFAM01.1	PPB	10/30/2024 20:03	SMPL
P4497-09	MBHDB4	836	0.2377 -		SFAM01.1	PPB	10/30/2024 20:05	SMPL
P4497-10	MBHDB5	815	0.2324 -		SFAM01.1	PPB	10/30/2024 20:07	SMPL

LB133224
INSTRUMENT ID : CV1

P4497-11	MBHDB6	1428	0.3869 -	SFAM01.1	PPB	10/30/2024 20:10 SMPL
P4497-12	MBHDB7	1263	0.3453 -	SFAM01.1	PPB	10/30/2024 20:12 SMPL
P4497-13	MBHDB8	874	0.2473 -	SFAM01.1	PPB	10/30/2024 20:14 SMPL
P4497-14	MBHDB9	1083	0.3 -	SFAM01.1	PPB	10/30/2024 20:16 SMPL
P4497-15	MBHDC0	973	0.2722 -	SFAM01.1	PPB	10/30/2024 20:19 SMPL
P4497-16	MBHDC1	602	0.1788 -	SFAM01.1	PPB	10/30/2024 20:21 SMPL
CCV001	CCV001	20489	5.1893 -	SFAM01.1	PPB	10/30/2024 20:23 SMPL
CCB001	CCB001	-71	0.0092 -	SFAM01.1	PPB	10/30/2024 20:25 SMPL
P4497-17	MBHDC2	1344	0.3657 -	SFAM01.1	PPB	10/30/2024 20:28 SMPL
P4497-18	MBHDC8	1249	0.3418 -	SFAM01.1	PPB	10/30/2024 20:30 SMPL
P4497-19	MBHDC9	1173	0.3226 -	SFAM01.1	PPB	10/30/2024 20:32 SMPL
P4497-20	MBHDD0	1848	0.4927 -	SFAM01.1	PPB	10/30/2024 20:35 SMPL
P4497-21	MBHDD0D	1736	0.4645 -	SFAM01.1	PPB	10/30/2024 20:37 SMPL
P4497-22	MBHDD0S	14491	3.6781 -	SFAM01.1	PPB	10/30/2024 20:41 SMPL
PB164551BL	PBS551	-80	0.0069 -	SFAM01.1	PPB	10/30/2024 20:43 SMPL
P4498-01	MBHDD1	483	0.1488 -	SFAM01.1	PPB	10/30/2024 20:46 SMPL
P4498-02	MBHDD2	720	0.2085 -	SFAM01.1	PPB	10/30/2024 20:48 SMPL
P4498-03	MBHDD8	284	0.0986 -	SFAM01.1	PPB	10/30/2024 20:50 SMPL
P4498-04	MBHDD9	602	0.1788 -	SFAM01.1	PPB	10/30/2024 20:53 SMPL
P4498-05	MBHDE0	1503	0.4058 -	SFAM01.1	PPB	10/30/2024 20:55 SMPL
P4498-06	MBHDE1	750	0.2161 -	SFAM01.1	PPB	10/30/2024 20:57 SMPL
P4498-07	MBHDE2	1287	0.3514 -	SFAM01.1	PPB	10/30/2024 20:59 SMPL
P4498-08	MBHDE8	1819	0.4854 -	SFAM01.1	PPB	10/30/2024 21:02 SMPL
P4498-09	MBHDE9	2218	0.5859 -	SFAM01.1	PPB	10/30/2024 21:04 SMPL
P4498-10	MBHDF0	2010	0.5335 -	SFAM01.1	PPB	10/30/2024 21:06 SMPL
P4498-11	MBHDF0D	1957	0.5202 -	SFAM01.1	PPB	10/30/2024 21:08 SMPL
P4498-12	MBHDF0S	13305	3.3793 -	SFAM01.1	PPB	10/30/2024 21:11 SMPL
P4498-13	MBHDF1	1271	0.3473 -	SFAM01.1	PPB	10/30/2024 21:13 SMPL
CCV002	CCV002	20308	5.1437 -	SFAM01.1	PPB	10/30/2024 21:15 SMPL
CCB002	CCB002	-74	0.0085 -	SFAM01.1	PPB	10/30/2024 21:18 SMPL
P4498-14	MBHDF2	1669	0.4476 -	SFAM01.1	PPB	10/30/2024 21:20 SMPL
P4498-15	MBHCH8	1181	0.3246 -	SFAM01.1	PPB	10/30/2024 21:22 SMPL
P4498-16	MBHCH9	1438	0.3894 -	SFAM01.1	PPB	10/30/2024 21:24 SMPL
P4498-17	MBHCJ0	1485	0.4012 -	SFAM01.1	PPB	10/30/2024 21:27 SMPL
P4498-18	MBHCJ1	1222	0.335 -	SFAM01.1	PPB	10/30/2024 21:29 SMPL
P4498-19	MBHCJ2	1517	0.4093 -	SFAM01.1	PPB	10/30/2024 21:31 SMPL
P4498-20	MBHCJ3	1272	0.3476 -	SFAM01.1	PPB	10/30/2024 21:34 SMPL
P4498-21	MBHCJ5	2020	0.536 -	SFAM01.1	PPB	10/30/2024 21:36 SMPL
P4498-22	MBHCJ6	1151	0.3171 -	SFAM01.1	PPB	10/30/2024 21:38 SMPL
CCV003	CCV003	20286	5.1381 -	SFAM01.1	PPB	10/30/2024 21:40 SMPL
CCB003	CCB003	-90	0.0044 -	SFAM01.1	PPB	10/30/2024 21:43 SMPL