

LB 133 227

7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5042E

A= 0.0000e+000

B= 2.0219e-004

C= 1.0513e-002

Rho= 0.9989018

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.047	0.047	179	0.000	179					7
0.2	0.200	0.214	0.014	1007	0.0 %	1007					7
2.5	2.500	2.679	0.179	13199	0.0 %	13199					7
5.0	5.000	4.738	-0.262	23382	0.0 %	23382					-5
7.5	7.500	7.316	-0.184	36133	0.0 %	36133					-2
10.0	10.000	10.206	0.206	50426	0.0 %	50426					2

LB33227 INSTRUMENT ID : CV

Method: 7471B

Operator: Admin

Date of Analysis: 31 Oct 2024 09:52:15

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	179	-	0.0000	7471B	PPB	31 Oct 2024 10:01:16	S	Std
0.2 - 1	502	1007	-	0.2000	7471B	PPB	31 Oct 2024 10:03:33	S	Std
2.5 - 1	52.5	13199	-	2.5000	7471B	PPB	31 Oct 2024 10:05:50	S	Std
5.0 - 1	55	23382	-	5.0000	7471B	PPB	31 Oct 2024 10:08:06	S	Std
7.5 - 1	57.5	36133	-	7.5000	7471B	PPB	31 Oct 2024 10:10:25	S	Std
10.0 - 1	520	50426	-	10.0000	7471B	PPB	31 Oct 2024 10:18:13	S	Std
ICV61 - 1	ICV61	18604	3.7720	-	7471B	PPB	31 Oct 2024 10:21:00	U	SMPL
ICB61 - 1	ICB61	-150	-0.0198	-	7471B	PPB	31 Oct 2024 10:23:16	U	SMPL
CCV04 - 1	CCV04	24081	4.8793	-	7471B	PPB	31 Oct 2024 10:25:35	U	SMPL
CCB04 - 1	CCB04	-104	-0.0105	-	7471B	PPB	31 Oct 2024 10:27:50	U	SMPL
CRA - 1	CRA	1070	0.2269	-	7471B	PPB	31 Oct 2024 10:30:07	U	SMPL
HighStd - 1	HighStd	50688	10.2589	-	7471B	PPB	31 Oct 2024 10:32:22	U	SMPL
ChkStd - 1	ChkStd	33977	6.8802	-	7471B	PPB	31 Oct 2024 10:34:37	U	SMPL
PB164552BL - 1	PBS	-351	-0.0605	-	7471B	PPB	31 Oct 2024 10:36:59	U	SMPL
PB164552BS - 1	LCSS	18082	3.6664	-	7471B	PPB	31 Oct 2024 10:39:19	U	SMPL
P4612-03 - 1	MOO-24-00335	971	0.2068	-	7471B	PPB	31 Oct 2024 10:46:57	U	SMPL
P4613-01 - 1	ARS20-0001	3415	0.7010	-	7471B	PPB	31 Oct 2024 10:49:14	U	SMPL
P4614-01 - 1	NB-07-102924	1858	0.3862	-	7471B	PPB	31 Oct 2024 10:51:33	U	SMPL
P4616-01 - 1	BP-F10	503	0.1122	-	7471B	PPB	31 Oct 2024 10:53:50	U	SMPL
CCV05 - 1	CCV05	24284	4.9204	-	7471B	PPB	31 Oct 2024 10:56:06	U	SMPL
CCB05 - 1	CCB05	-62	-0.0020	-	7471B	PPB	31 Oct 2024 11:01:14	U	SMPL
P4616-05 - 1	BP-F9-MOVED	611	0.1340	-	7471B	PPB	31 Oct 2024 11:03:28	U	SMPL
P4617-01 - 1	RO-SOIL	7483	1.5235	-	7471B	PPB	31 Oct 2024 11:05:44	U	SMPL
P4618-01 - 1	HR-01-102924	2056	0.4262	-	7471B	PPB	31 Oct 2024 11:07:59	U	SMPL
P4618-01DUP - 1	HR-01-102924DUP	2385	0.4927	-	7471B	PPB	31 Oct 2024 11:10:17	U	SMPL
P4618-01MS - 1	HR-01-102924MS	19791	4.0120	-	7471B	PPB	31 Oct 2024 11:12:33	U	SMPL
P4618-01MSD - 1	HR-01-102924MSD	21551	4.3678	-	7471B	PPB	31 Oct 2024 11:14:50	U	SMPL
P4618-03 - 1	HR-04-102924	6175	1.2590	-	7471B	PPB	31 Oct 2024 11:17:09	U	SMPL
P4619-01 - 1	SU-03-102924	986	0.2099	-	7471B	PPB	31 Oct 2024 11:19:28	U	SMPL
P4639-01 - 1	EO-01-103024	3794	0.7776	-	7471B	PPB	31 Oct 2024 11:21:47	U	SMPL
P4639-03 - 1	EO-02-103024	4036	0.8265	-	7471B	PPB	31 Oct 2024 11:24:02	U	SMPL
CCV06 - 1	CCV06	24426	4.9491	-	7471B	PPB	31 Oct 2024 11:26:19	U	SMPL
CCB06 - 1	CCB06	114	0.0336	-	7471B	PPB	31 Oct 2024 11:31:45	U	SMPL
P4640-01 - 1	MH-3	10295	2.0920	-	7471B	PPB	31 Oct 2024 11:34:00	U	SMPL
P4495-10 - 1	PT-MET-SOIL	775962	HIGH	-	7471B	PPB	31 Oct 2024 11:38:05	X	
P4618-01LX5 - 1		340	0.0793	-	7471B	PPB	31 Oct 2024 11:43:46	U	SMPL
P4618-01A - 1		20927	4.2417	-	7471B	PPB	31 Oct 2024 11:46:01	U	SMPL
P4495-10DLX10 - 1	PT-MET-SOIL	178693	36.1397	-	7471B	PPB	31 Oct 2024 11:53:11	U	SMPL
P4495-10DL2X50 - 1	PT-MET-SOIL	33253	6.7338	-	7471B	PPB	31 Oct 2024 12:00:37	U	SMPL
CCV07 - 1	CCV07	24337	4.9311	-	7471B	PPB	31 Oct 2024 12:02:57	U	SMPL
CCB07 - 1	CCB07	-113	-0.0123	-	7471B	PPB	31 Oct 2024 12:05:16	U	SMPL