

LB133232


7470A

INSTRUMENT 19: CV2

Linear

 μ Abs.:

36230



A= 0.0000e+000

B= 2.7770e-004

C= -2.2868e-002

Rho= 0.9993967

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	2/01
0.0	0.000	-0.028	-0.028	-19	0.000	-19	0				1
0.2	0.200	0.192	-0.008	773	0.0 %	773					4
2.5	2.500	2.708	0.208	9833	0.0 %	9833					8
5.0	5.000	4.770	-0.230	17261	0.0 %	17261					5
7.5	7.500	7.520	0.020	27162	0.0 %	27162					0
10.0	10.000	10.038	0.038	36230	0.0 %	36230					0

LB133232

Method: 7470A

Operator: Admin

Date of Analysis: 31 Oct 2024 13:47:27

Sample ID	Extended ID	Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.0000	7470A	PPB	31 Oct 2024 13:50:31	S	Std
0.2 - 1	50.2	773	-	0.2000	7470A	PPB	31 Oct 2024 13:52:48	S	Std
2.5 - 1	52.5	9833	-	2.5000	7470A	PPB	31 Oct 2024 13:55:05	S	Std
5.0 - 1	55	17261	-	5.0000	7470A	PPB	31 Oct 2024 14:00:02	S	Std
7.5 - 1	57.5	27162	-	7.5000	7470A	PPB	31 Oct 2024 14:02:18	S	Std
10.0 - 1	57.5	36230	-	10.0000	7470A	PPB	31 Oct 2024 14:04:35	S	Std
ICV62 - 1	ICV62	13590	3.7511	-	7470A	PPB	31 Oct 2024 14:07:24	U	SMPL
ICB62 - 1	ICB62	-275	-0.0992	-	7470A	PPB	31 Oct 2024 14:09:39	U	SMPL
CCV08 - 1	CCV08	17245	4.7660	-	7470A	PPB	31 Oct 2024 14:11:55	U	SMPL
CCB08 - 1	CCB08	-217	-0.0831	-	7470A	PPB	31 Oct 2024 14:14:09	U	SMPL
CRA - 1	CRA	706	0.1732	-	7470A	PPB	31 Oct 2024 14:16:26	U	SMPL
HighStd - 1	HighStd	35607	9.8651	-	7470A	PPB	31 Oct 2024 14:18:41	U	SMPL
ChkStd - 1	ChkStd	23255	6.4350	-	7470A	PPB	31 Oct 2024 14:20:56	U	SMPL
PB164565BL - 1	PBW	-80	-0.0451	-	7470A	PPB	31 Oct 2024 14:26:32	U	SMPL
PB164565BS - 1	LCSW	13282	3.6655	-	7470A	PPB	31 Oct 2024 14:28:48	U	SMPL
P3390-09 - 1	MDL-WATER-03-QT3-2024	315	0.0646	-	7470A	PPB	31 Oct 2024 14:35:08	U	SMPL
P4596-07 - 1	TOTE-WATER-COMP	1336	0.3481	-	7470A	PPB	31 Oct 2024 14:41:52	U	SMPL
P4596-07DUP - 1	TOTE-WATER-COMPDUP	1116	0.2870	-	7470A	PPB	31 Oct 2024 14:44:09	U	SMPL
P4596-07MS - 1	TOTE-WATER-COMPMS	14428	3.9838	-	7470A	PPB	31 Oct 2024 14:46:26	U	SMPL
P4596-07MSD - 1	TOTE-WATER-COMPMSD	13945	3.8496	-	7470A	PPB	31 Oct 2024 14:48:41	U	SMPL
CCV09 - 1	CCV09	17370	4.8008	-	7470A	PPB	31 Oct 2024 14:50:58	U	SMPL
CCB09 - 1	CCB09	-106	-0.0523	-	7470A	PPB	31 Oct 2024 14:53:15	U	SMPL
P3390-07 - 1		286	0.0566	-	7470A	PPB	31 Oct 2024 14:58:07	U	SMPL
P3390-07RE - 1		551	0.1301	-	7470A	PPB	31 Oct 2024 15:00:22	U	SMPL
P4596-07LX5 - 1		305	0.0618	-	7470A	PPB	31 Oct 2024 15:02:37	U	SMPL
P4596-07A - 1		14225	3.9274	-	7470A	PPB	31 Oct 2024 15:04:53	U	SMPL
CCV10 - 1	CCV10	17356	4.7969	-	7470A	PPB	31 Oct 2024 15:07:08	U	SMPL
CCB10 - 1	CCB10	-70	-0.0423	-	7470A	PPB	31 Oct 2024 15:09:26	U	SMPL