

LB133246

7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

3472E

A= 0.0000e+000

B= 2.8034e-004

C= 4.8970e-002

Rho= 0.9990589

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/a
0.0	0.000	0.022	0.022	-95	0.000	-95					
0.2	0.200	0.193	-0.007	512	0.0 %	512					
2.5	2.500	2.414	-0.086	8438	0.0 %	8438					
5.0	5.000	4.972	-0.028	17561	0.0 %	17561					
7.5	7.500	7.814	0.314	27699	0.0 %	27699					
10.0	10.000	9.785	-0.215	34728	0.0 %	34728					

o/a

-4

-3

-1

4

-2

LB133246 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Nov 2024 09:54:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-95	-	0.0000	7470A	PPB	01 Nov 2024 10:00:11	S	Std
0.2 - 1	50.2	512	-	0.2000	7470A	PPB	01 Nov 2024 10:02:27	S	Std
2.5 - 1	52.5	8438	-	2.5000	7470A	PPB	01 Nov 2024 10:04:44	S	Std
5.0 - 1	55	17561	-	5.0000	7470A	PPB	01 Nov 2024 10:07:00	S	Std
7.5 - 1	57.5	27699	-	7.5000	7470A	PPB	01 Nov 2024 10:09:17	S	Std
10.0 - 1	510	34728	-	10.0000	7470A	PPB	01 Nov 2024 10:14:20	S	Std
ICV65 - 1	ICV65	13822	3.9238	-7470A	PPB	01 Nov 2024 10:17:09	U	SMPL	
ICB65 - 1	ICB65	-439	-0.0741	-7470A	PPB	01 Nov 2024 10:19:25	U	SMPL	
CCV13 - 1	CCV13	16713	4.7343	-7470A	PPB	01 Nov 2024 10:21:42	U	SMPL	
CCB13 - 1	CCB13	-371	-0.0550	-7470A	PPB	01 Nov 2024 10:23:57	U	SMPL	
CRA - 1	CRA	546	0.2020	-7470A	PPB	01 Nov 2024 10:26:14	U	SMPL	
HighStd - 1	HighStd	34033	9.5898	-7470A	PPB	01 Nov 2024 10:28:28	U	SMPL	
ChkStd - 1	ChkStd	22770	6.4323	-7470A	PPB	01 Nov 2024 10:30:43	U	SMPL	
PB164580BL - 1	PBW	-513	-0.0948	-7470A	PPB	01 Nov 2024 10:33:01	U	SMPL	
PB164580BS - 1	LCSW	12890	3.6625	-7470A	PPB	01 Nov 2024 10:35:18	U	SMPL	
P4611-03 - 1	TP-1	-331	-0.0438	-7470A	PPB	01 Nov 2024 10:37:35	U	SMPL	
P4611-06 - 1	TP-2	2	0.0495	-7470A	PPB	01 Nov 2024 10:39:53	U	SMPL	
P4611-09 - 1	TP-3	-122	0.0148	-7470A	PPB	01 Nov 2024 10:42:10	U	SMPL	
P4611-12 - 1	TP-4	-103	0.0201	-7470A	PPB	01 Nov 2024 10:44:27	U	SMPL	
P4611-15 - 1	TP-5	-23	0.0425	-7470A	PPB	01 Nov 2024 10:46:42	U	SMPL	
CCV14 - 1	CCV14	17103	4.8436	-7470A	PPB	01 Nov 2024 10:48:57	U	SMPL	
CCB14 - 1	CCB14	-254	-0.0222	-7470A	PPB	01 Nov 2024 10:51:12	U	SMPL	
P4611-18 - 1	TP-6	19	0.0543	-7470A	PPB	01 Nov 2024 10:53:28	U	SMPL	
P4612-04 - 1	MOO-24-00335	-148	0.0075	-7470A	PPB	01 Nov 2024 10:55:43	U	SMPL	
P4613-02 - 1	ARS20-0001	-128	0.0131	-7470A	PPB	01 Nov 2024 10:57:58	U	SMPL	
P4616-04 - 1	BP-F10	-101	0.0207	-7470A	PPB	01 Nov 2024 11:00:14	U	SMPL	
P4616-08 - 1	BP-F9-MOVED	-99	0.0212	-7470A	PPB	01 Nov 2024 11:02:29	U	SMPL	
P4617-04 - 1	CONCRETE-PILE	-75	0.0279	-7470A	PPB	01 Nov 2024 11:04:45	U	SMPL	
P4617-04DUP - 1	CONCRETE-PILEDUP	-14	0.0450	-7470A	PPB	01 Nov 2024 11:07:02	U	SMPL	
P4617-04MS - 1	CONCRETE-PILEMS	13922	3.9519	-7470A	PPB	01 Nov 2024 11:09:19	U	SMPL	
P4617-04MSD - 1	CONCRETE-PILEMSD	13604	3.8627	-7470A	PPB	01 Nov 2024 11:11:35	U	SMPL	
CCV15 - 1	CCV15	17727	5.0185	-7470A	PPB	01 Nov 2024 11:13:52	U	SMPL	
CCB15 - 1	CCB15	-435	-0.0730	-7470A	PPB	01 Nov 2024 11:16:09	U	SMPL	
PB164522TB - 1	PB164522TB	-187	-0.0035	-7470A	PPB	01 Nov 2024 11:18:26	U	SMPL	
P4617-04LX5 - 1		-34	0.0394	-7470A	PPB	01 Nov 2024 11:20:42	U	SMPL	
P4617-04A - 1		13783	3.9129	-7470A	PPB	01 Nov 2024 11:22:57	U	SMPL	
CCV16 - 1	CCV16	16960	4.8035	-7470A	PPB	01 Nov 2024 11:30:03	U	SMPL	
CCB16 - 1	CCB16	-257	-0.0231	-7470A	PPB	01 Nov 2024 11:32:18	U	SMPL	