

LB133253


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

32270



A= 0.0000e+000

B= 3.0507e-004

C= -3.0828e-003

Rho= 0.9996691

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	-0.021	-0.021	-58	0.000	-58					-18
0.05	0.050	0.041	-0.009	143	0.0 %	143	0				-19
0.2	0.200	0.163	-0.037	545	0.0 %	545					-2
2.5	2.500	2.461	-0.039	8076	0.0 %	8076					3
5.0	5.000	5.144	0.144	16871	0.0 %	16871					2
7.5	7.500	7.621	0.121	24991	0.0 %	24991					-2
10.0	10.000	9.842	-0.158	32270	0.0 %	32270					

LB133252 INSTRUMERNT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 Nov 2024 15:56:25

Sample ID	Extended ID	µ Abs.	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	-58	-	0.0000	245.1	PPB	01 Nov 2024 15:58:21	S	Std
0.05 - 1	50.05	143	-	0.0500	245.1	PPB	01 Nov 2024 16:00:38	S	Std
0.2 - 1	50.2	545	-	0.2000	245.1	PPB	01 Nov 2024 16:02:55	S	Std
2.5 - 1	82.5	8075	-	2.5000	245.1	PPB	01 Nov 2024 16:05:11	S	Std
5.0 - 1	85	16871	-	5.0000	245.1	PPB	01 Nov 2024 16:07:27	S	Std
7.5 - 1	375	24991	-	7.5000	245.1	PPB	01 Nov 2024 16:09:44	S	Std
10.0 - 1	375	32270	-	10.0000	245.1	PPB	01 Nov 2024 16:18:05	S	Std
ICV67 - 1	ICV67	12700	3.8713	-245.1		PPB	01 Nov 2024 16:23:30	U	SMPL
ICB67 - 1	ICB67	-279	-0.0882	-245.1		PPB	01 Nov 2024 16:25:45	U	SMPL
CCV19 - 1	CCV19	16480	5.0245	-245.1		PPB	01 Nov 2024 16:28:02	U	SMPL
CCB19 - 1	CCB19	-243	-0.0772	-245.1		PPB	01 Nov 2024 16:30:17	U	SMPL
CRA - 1	CRA	174	0.0500	-245.1		PPB	01 Nov 2024 16:39:39	U	SMPL
HighStd - 1	HighStd	31480	9.6006	-245.1		PPB	01 Nov 2024 16:41:54	U	SMPL
ChkStd - 1	ChkStd	20801	6.3427	-245.1		PPB	01 Nov 2024 16:44:08	U	SMPL
PB164611BL - 1	PBW	-295	-0.0931	-245.1		PPB	01 Nov 2024 16:46:27	U	SMPL
PB164611BS - 1	LCSW	14411	4.3933	-245.1		PPB	01 Nov 2024 16:48:45	U	SMPL
P3390-10 - 1	LOD-MDL-WATER-04-QT3-2024	125	0.0351	-245.1		PPB	01 Nov 2024 16:58:51	U	SMPL
P3390-11 - 1	LOQ-WATER-05-QT3-2024	562	0.1684	-245.1		PPB	01 Nov 2024 17:03:07	U	SMPL
P4632-02 - 1	EFFLUENT-METALS	250	0.0732	-245.1		PPB	01 Nov 2024 17:05:24	U	SMPL
P4632-02DUP - 1	EFFLUENT-METALS DUP	49	0.0119	-245.1		PPB	01 Nov 2024 17:07:41	U	SMPL
P4632-02MS - 1	EFFLUENT-METALS MS	10742	3.2740	-245.1		PPB	01 Nov 2024 17:09:56	U	SMPL
CCV20 - 1	CCV20	15846	4.8311	-245.1		PPB	01 Nov 2024 17:12:11	U	SMPL
CCB20 - 1	CCB20	-197	-0.0632	-245.1		PPB	01 Nov 2024 17:14:27	U	SMPL
P4632-02MSD - 1	EFFLUENT-METALS MSD	11293	3.4421	-245.1		PPB	01 Nov 2024 17:16:44	U	SMPL
P4632-02LX5 - 1		-181	-0.0583	-245.1		PPB	01 Nov 2024 17:18:59	U	SMPL
P4632-02A - 1		12192	3.7164	-245.1		PPB	01 Nov 2024 17:21:15	U	SMPL
CCV21 - 1	CCV21	15857	4.8344	-245.1		PPB	01 Nov 2024 17:23:30	U	SMPL
CCB21 - 1	CCB21	-274	-0.0867	-245.1		PPB	01 Nov 2024 17:25:48	U	SMPL