

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133270

Ambient Barometric Pressure (mmHg): 781.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.1	11/04/2024	09:20
2	P4661-03		1	12	100.00	>212.0	>212.0	11/04/2024	09:50
3	P4661-03DUP		1	12	100.00	>212.0	>212.0	11/04/2024	10:20
4	P4662-04		1	11	100.00	>212.0	>212.0	11/04/2024	10:50
5	P4684-03		1	14	100.00	>212.0	>212.0	11/04/2024	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133270

WorkList Name : fp-11-01

WorkList ID : 185036

Department : Wet-Chemistry

Date : 11-01-2024 15:48:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4661-03	102324-A	Water	Flash Point	Cool 4 deg C	PSEG03	K51	10/31/2024	1010B
P4662-04	102524-D	Water	Flash Point	Cool 4 deg C	PSEG03	K61	10/31/2024	1010B
P4684-03	MECHANIC-ST-OIL-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	L13	11/01/2024	1010B

Date/Time 11/03/2024 09:10
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: JWC

Date/Time 11/03/2024 12:00
Raw Sample Received by: JWC
Raw Sample Relinquished by: RM CWC