

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB133271

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/04/2024	09:30
2	CAL2	1	Water	NA	NA	20.3	7.00	11/04/2024	09:31
3	CAL3	1	Water	NA	NA	20.2	10.02	11/04/2024	09:33
4	ICV	1	Water	NA	NA	20.3	7.01	11/04/2024	09:35
5	CCV1	1	Water	NA	NA	20.2	2.01	11/04/2024	09:37
6	P4660-03	1	Solid	20.02	20	24.2	11.87	11/04/2024	09:50
7	P4660-03DUP	1	Solid	20.03	20	24.3	11.88	11/04/2024	09:51
8	P4660-07	1	Solid	20.01	20	24.3	10.96	11/04/2024	10:00
9	P4660-11	1	Solid	20.03	20	24.3	8.81	11/04/2024	10:10
10	P4667-04	1	Solid	20.02	20	24.1	8.42	11/04/2024	10:14
11	P4667-08	1	Solid	20.03	20	24.5	8.12	11/04/2024	10:15
12	P4667-12	1	Solid	20.02	20	23.9	8.14	11/04/2024	10:20
13	P4667-16	1	Solid	20.03	20	24.2	8.77	11/04/2024	10:24
14	P4679-04	1	Solid	20.03	20	23.9	8.73	11/04/2024	10:30
15	P4680-04	1	Solid	20.04	20	24.3	9.38	11/04/2024	10:35
16	CCV2	1	Water	NA	NA	20.3	12.02	11/04/2024	10:40
17	P4680-08	1	Solid	20.02	20	23.9	8.65	11/04/2024	10:45
18	P4682-02	1	Solid	20.03	20	24.3	5.31	11/04/2024	10:50
19	CCV3	1	Water	NA	NA	20.3	2.01	11/04/2024	10:55

WORKLIST(Hardcopy Internal Chain)

WB 133271

WorkList Name : corrosivity p4679 WorkList ID : 185085 Department : Wet-Chemistry Date : 11-04-2024 09:19:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4660-03	WC-TA2-01-C	Solid	Corrosivity	Cool 4 deg C	ENTAO5	K41	10/30/2024	9045D
P4660-07	WC-WOOD-01-C	Solid	Corrosivity	Cool 4 deg C	ENTAO5	K41	10/31/2024	9045D
P4660-11	WC-CONCRETE-01-C	Solid	Corrosivity	Cool 4 deg C	ENTAO5	K41	10/31/2024	9045D
P4667-04	BP-F-6	Solid	Corrosivity	Cool 4 deg C	PSEG03	K51	10/31/2024	9045D
P4667-08	BP-F-5	Solid	Corrosivity	Cool 4 deg C	PSEG03	K51	10/31/2024	9045D
P4667-12	BP-F-10	Solid	Corrosivity	Cool 4 deg C	PSEG03	K51	10/31/2024	9045D
P4667-16	BP-F-7	Solid	Corrosivity	Cool 4 deg C	PSEG03	K51	10/31/2024	9045D
P4679-04	MH-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	K51	10/31/2024	9045D
P4680-04	BP-F26	Solid	Corrosivity	Cool 4 deg C	PSEG03	L11	11/01/2024	9045D
P4680-08	BP-F25	Solid	Corrosivity	Cool 4 deg C	PSEG03	L12	11/01/2024	9045D
P4682-02	BELL-SHOP-RAGS	Solid	Corrosivity	Cool 4 deg C	PSEG03	L12	11/01/2024	9045D
					PSEG03	L13	11/01/2024	9045D

Date/Time 11/04/24 09:25
Raw Sample Received by: SO gwc,
Raw Sample Relinquished by: [Signature]

Date/Time 11/04/24 13:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]