

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133272

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/04/2024	09:30
2	CAL2	1	Water	NA	NA	20.3	7.00	11/04/2024	09:31
3	CAL3	1	Water	NA	NA	20.2	10.02	11/04/2024	09:33
4	ICV	1	Water	NA	NA	20.3	7.01	11/04/2024	09:35
5	CCV1	1	Water	NA	NA	20.3	2.02	11/04/2024	11:00
6	P4661-03	1	Water	NA	NA	21.9	6.26	11/04/2024	11:10
7	P4662-04	1	Water	NA	NA	22.2	5.17	11/04/2024	11:15
8	P4684-03	1	Water	NA	NA	22.6	5.72	11/04/2024	11:25
9	P4684-03DUP	1	Water	NA	NA	22.7	5.73	11/04/2024	11:26
10	CCV2	1	Water	NA	NA	20.3	12.02	11/04/2024	11:30

WORKLIST(Hardcopy Internal Chain)

UP 133272

WorkList Name : ph water p4684 WorkList ID : 185086 Department : Wet-Chemistry Date : 11-04-2024 09:20:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4661-03	102324-A	Water	pH	Cool 4 deg C	PSEG03	K51	10/31/2024	9040C
P4662-04	102524-D	Water	pH	Cool 4 deg C	PSEG03	K61	10/31/2024	9040C
P4684-03	MECHANIC-ST-OIL-WATER	Water	pH	Cool 4 deg C	PSEG03	L13	11/01/2024	9040C

Date/Time 11/04/24 09:25
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11/04/24 11:50
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]