

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB133284

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/04/2024	15:50
2	CAL2	1	Water	NA	NA	20.3	7.00	11/04/2024	15:51
3	CAL3	1	Water	NA	NA	20.3	10.02	11/04/2024	15:52
4	ICV	1	Water	NA	NA	20.3	7.01	11/04/2024	15:55
5	CCV1	1	Water	NA	NA	20.3	2.01	11/04/2024	16:00
6	P4693-04	1	Solid	20.02	20	22.9	9.19	11/04/2024	16:10
7	P4693-08	1	Solid	20.03	20	22.7	9.72	11/04/2024	16:11
8	P4694-04	1	Solid	20.02	20	22.5	10.00	11/04/2024	16:15
9	P4694-08	1	Solid	20.03	20	22.4	9.80	11/04/2024	16:22
10	P4695-04	1	Solid	20.03	20	22.2	7.84	11/04/2024	16:25
11	P4700-04	1	Solid	20.03	20	24.4	8.66	11/04/2024	16:30
12	P4701-04	1	Solid	20.02	20	22.8	8.29	11/04/2024	16:33
13	P4701-08	1	Solid	20.03	20	23.1	8.78	11/04/2024	16:37
14	P4711-05	1	Solid	20.02	20	20.8	6.24	11/04/2024	16:44
15	P4711-10	1	Solid	20.03	20	22.4	5.85	11/04/2024	16:50
16	CCV2	1	Water	NA	NA	20.2	12.02	11/04/2024	16:55
17	P4711-10DUP	1	Solid	20.04	20	22.5	5.86	11/04/2024	16:56
18	CCV3	1	Water	NA	NA	20.3	2.01	11/04/2024	17:00

WORKLIST(Hardcopy Internal Chain)

VP 133284

WorkList Name : corrosivity p4695 WorkList ID : 185112 Department : Wet-Chemistry Date : 11-04-2024 12:04:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4693-04	BP-G5-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/02/2024	9045D
P4693-08	BP-G4-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/02/2024	9045D
P4694-04	Z-03A	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/03/2024	9045D
P4694-08	Z-04	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/03/2024	9045D
P4695-04	Z-01	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/01/2024	9045D
P4700-04	MH-8	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	11/04/2024	9045D
P4701-04	BP-F3	Solid	Corrosivity	Cool 4 deg C	PSEG03	K23	11/04/2024	9045D
P4701-08	BP-F4	Solid	Corrosivity	Cool 4 deg C	PSEG03	K23	11/04/2024	9045D
P4711-05	CF-613-COMP-16	Solid	Corrosivity	Cool 4 deg C	PSEG03	K51	11/04/2024	9045D
P4711-10	CF-613-COMP-17	Solid	Corrosivity	Cool 4 deg C	PSEG03	K51	11/04/2024	9045D

Date/Time 11/04/24 15:30
 Raw Sample Received by: JD WOC
 Raw Sample Relinquished by: JD WOC

Date/Time 11/04/24 18:30
 Raw Sample Received by: JD WOC
 Raw Sample Relinquished by: JD WOC