

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB133285

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/05/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	11/05/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.03	11/05/2024	08:12
4	ICV	1	Water	NA	NA	20.3	7.02	11/05/2024	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	11/05/2024	08:20
6	P4702-01	1	Water	NA	NA	20.4	8.59	11/05/2024	08:25
7	P4702-01DUP	1	Water	NA	NA	20.5	8.60	11/05/2024	08:26
8	P4704-01	1	Water	NA	NA	20.7	4.82	11/05/2024	08:33
9	P4704-02	1	Water	NA	NA	20.4	7.11	11/05/2024	08:37
10	P4704-03	1	Water	NA	NA	20.4	7.07	11/05/2024	08:44
11	CCV2	1	Water	NA	NA	20.3	12.02	11/05/2024	08:45

WORKLIST(Hardcopy Internal Chain)

UB 133285

WorkList Name : ph water p4702 WorkList ID : 185119 Department : Wet-Chemistry Date : 11-05-2024 07:46:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4702-01	C TOTE-1	Water	pH	Cool 4 deg C	PSEG03	L23	11/04/2024	9040C
P4704-01	A NWB-2111	Water	pH	Cool 4 deg C	PSEG03	L23	11/04/2024	9040C
P4704-02	C NWB-2056	Water	pH	Cool 4 deg C	PSEG03	L23	11/04/2024	9040C
P4704-03	C NWB-2108	Water	pH	Cool 4 deg C	PSEG03	L23	11/04/2024	9040C

Date/Time 11/05/24 08:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11/05/24 11:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]