

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133286

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/05/2024	09:10
2	CAL2	1	Water	NA	NA	20.3	7.00	11/05/2024	09:11
3	CAL3	1	Water	NA	NA	20.3	10.02	11/05/2024	09:12
4	ICV	1	Water	NA	NA	20.3	7.00	11/05/2024	09:15
5	CCV1	1	Water	NA	NA	20.3	2.02	11/05/2024	09:20
6	P4692-01	1	Solid	20.02	20	23.5	6.01	11/05/2024	09:30
7	P4692-01DUP	1	Solid	20.03	20	23.6	6.02	11/05/2024	09:31
8	P4692-02	1	Solid	20.02	20	22.7	6.24	11/05/2024	09:35
9	CCV2	1	Water	NA	NA	20.3	12.02	11/05/2024	09:37

WORKLIST(Hardcopy Internal Chain)

JB 133286

WorkList Name : ph 4692

WorkList ID : 185120

Department : Wet-Chemistry

Date : 11-05-2024 07:47:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4692-01	☒ TAPLPR-SED02-103124-00-T2	Solid	pH	Cool 4 deg C	WEST04	L21	10/31/2024	9045D
P4692-02	☒ TAPLPR-SED01-103124-00-T2	Solid	pH	Cool 4 deg C	WEST04	L21	10/31/2024	9045D

Date/Time 11/05/24 08:00

Raw Sample Received by: APG

Raw Sample Relinquished by: APG

Date/Time 11/05/24

Raw Sample Received by: APG

Raw Sample Relinquished by: APG