

66133302

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

11/5/2024 13:12

Test: Ammonia-N

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.038	0.0	0.150	
ICB1	0.007	0.0	0.032	
CCV1	1.022	0.0	0.148	
CCB1	0.010	0.0	0.033	
RL CHECK	0.099	0.0	0.043	
PB164458BL	0.010	0.0	0.032	
PB164458BS	1.039	0.0	0.150	
P4495-07	27.160	0.0	3.123	
P4675-01	0.030	0.0	0.035	
P4675-01DUP	0.013	0.0	0.033	
P4675-01MS	1.083	0.0	0.155	
P4675-01MSD	0.999	0.0	0.145	
P4675-02	0.049	0.0	0.037	
P4675-03	0.011	0.0	0.033	
CCV2	1.028	0.0	0.148	
CCB2	0.022	0.0	0.034	
P4675-04	0.037	0.0	0.036	
P4675-05	0.019	0.0	0.033	
P4675-06	0.025	0.0	0.034	
CCV3	1.042	0.0	0.150	
CCB3	0.033	0.0	0.035	
P4495-07DLX40	1.087	0.0	0.155	
CCV4	1.030	0.0	0.149	
CCB4	0.005	0.0	0.032	

99/(50-150) 11/05/2024 RM
Init abs., Test limit hig

N	24
Mean	1.537
SD	5.4800
CV%	356.47

Aquakem v. 7.2AQ1

Results from time period:

Tue Nov 05 10:50:20 2024

Tue Nov 05 12:29:06 2024

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Ammonia-† P		0.0123	mg/l	11/5/2024 8:44:23	
0.1PPM	A	Ammonia-† P		0.1131	mg/l	11/5/2024 8:44:24	
0.2PPM	A	Ammonia-† P		0.1921	mg/l	11/5/2024 8:44:25	
0.4PPM	A	Ammonia-† P		0.3915	mg/l	11/5/2024 8:44:26	
1.0PPM	A	Ammonia-† P		1.0168	mg/l	11/5/2024 8:44:27	
1.3PPM	A	Ammonia-† P		1.2758	mg/l	11/5/2024 8:44:28	
2.0PPM	A	Ammonia-† P		2.0318	mg/l	11/5/2024 8:44:29	
ICV1	S	Ammonia-† P		1.0377	mg/l	11/5/2024 10:50:21	
ICB1	S	Ammonia-† P		0.0074	mg/l	11/5/2024 10:50:22	
CCV1	S	Ammonia-† P		1.0221	mg/l	11/5/2024 10:50:25	
CCB1	S	Ammonia-† P		0.0104	mg/l	11/5/2024 10:50:26	
RL CHECK	S	Ammonia-† P		0.0993	mg/l	11/5/2024 10:50:29	
PB164458BL	S	Ammonia-† P		0.0097	mg/l	11/5/2024 10:50:31	
PB164458BS	S	Ammonia-† P		1.0387	mg/l	11/5/2024 11:01:04	
P4495-07	S	Ammonia-† P		27.1596	mg/l	11/5/2024 11:01:07	
P4675-01	S	Ammonia-† P		0.0295	mg/l	11/5/2024 11:11:51	
P4675-01DUP	S	Ammonia-† P		0.0131	mg/l	11/5/2024 11:11:53	
P4675-01MS	S	Ammonia-† P		1.0826	mg/l	11/5/2024 11:11:54	
P4675-01MSD	S	Ammonia-† P		0.9988	mg/l	11/5/2024 11:11:55	
P4675-02	S	Ammonia-† P		0.0488	mg/l	11/5/2024 11:11:58	
P4675-03	S	Ammonia-† P		0.0109	mg/l	11/5/2024 11:11:59	
CCV2	S	Ammonia-† P		1.0285	mg/l	11/5/2024 11:22:30	
CCB2	S	Ammonia-† P		0.0215	mg/l	11/5/2024 11:22:32	
P4675-04	S	Ammonia-† P		0.0367	mg/l	11/5/2024 11:22:33	
P4675-05	S	Ammonia-† P		0.0187	mg/l	11/5/2024 11:22:34	
P4675-06	S	Ammonia-† P		0.0247	mg/l	11/5/2024 11:22:35	
CCV3	S	Ammonia-† P		1.0421	mg/l	11/5/2024 11:22:40	
CCB3	S	Ammonia-† P		0.0329	mg/l	11/5/2024 11:27:59	
P4495-07DLX40	S	Ammonia-† P		1.0875	mg/l	11/5/2024 12:29:00	
CCV4	S	Ammonia-† P		1.0296	mg/l	11/5/2024 12:29:04	
CCB4	S	Ammonia-† P		0.0046	mg/l	11/5/2024 12:29:06	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

11/5/2024 8:46


Test Ammonia-N

Accepted 11/5/2024 8:46

Factor 8.784
Bias 0.031

Coeff. of det. 0.998499

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.033	0.0123	0.0000	13.1
2	NH3-2PPM	0.044	0.1131	0.1000	-4.0
3	NH3-2PPM	0.053	0.1921	0.2000	-2.1
4	NH3-2PPM	0.076	0.3915	0.4000	1.7
5	NH3-2PPM	0.147	1.0168	1.0000	-1.9
6	NH3-2PPM	0.177	1.2758	1.3333	1.6
7	NH3-2PPM	0.263	2.0318	2.0000	

11/05/2024
RM