


LB133311

7471B

INSTRUMENT ID: 142

Linear



A= 0.0000e+000

B= 2.2145e-004

C= -3.0643e-003

Rho= 0.9987440

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.017	-0.017	-64	0.000	-64					-19
0.2	0.200	0.163	-0.037	748	0.0 %	748					-4
2.5	2.500	2.398	-0.102	10843	0.0 %	10843					3
5.0	5.000	5.156	0.156	23298	0.0 %	23298					4
7.5	7.500	7.790	0.290	35192	0.0 %	35192					-3
10.0	10.000	9.710	-0.290	43859	0.0 %	43859					

LB133311 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Nov 2024 10:24:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-64	-	0.0000	7471B	PPB	06 Nov 2024 10:32:48	S	Std
0.2 - 1	50.2	748	-	0.2000	7471B	PPB	06 Nov 2024 10:35:05	S	Std
2.5 - 1	52.5	10843	-	2.5000	7471B	PPB	06 Nov 2024 10:37:22	S	Std
5.0 - 1	55	23298	-	5.0000	7471B	PPB	06 Nov 2024 10:39:39	S	Std
7.5 - 1	57.5	35192	-	7.5000	7471B	PPB	06 Nov 2024 10:41:58	S	Std
10.0 - 1	59	43859	-	10.0000	7471B	PPB	06 Nov 2024 10:49:18	S	Std
ICV70 - 1	ICV70	17897	3.9603	-	7471B	PPB	06 Nov 2024 10:52:10	U	SMPL
ICB70 - 1	ICB70	-268	-0.0624	-	7471B	PPB	06 Nov 2024 10:54:26	U	SMPL
CCV32 - 1	CCV32	24133	5.3413	-	7471B	PPB	06 Nov 2024 10:56:44	U	SMPL
CCB32 - 1	CCB32	-326	-0.0753	-	7471B	PPB	06 Nov 2024 10:58:58	U	SMPL
CRA - 1	CRA	977	0.2133	-	7471B	PPB	06 Nov 2024 11:04:01	U	SMPL
HighStd - 1	HighStd	44241	9.7943	-	7471B	PPB	06 Nov 2024 11:06:15	U	SMPL
ChkStd - 1	ChkStd	33165	7.3415	-	7471B	PPB	06 Nov 2024 11:08:30	U	SMPL
PB164691BL - 1	PBS	-323	-0.0746	-	7471B	PPB	06 Nov 2024 11:10:54	U	SMPL
PB164691BS - 1	LCSS	16960	3.7528	-	7471B	PPB	06 Nov 2024 11:13:16	U	SMPL
P4682-01 - 1	BELL-SHOP-RAGS	1223	0.2678	-	7471B	PPB	06 Nov 2024 11:15:32	U	SMPL
P4685-01 - 1	OK-01-11012024	3188	0.7029	-	7471B	PPB	06 Nov 2024 11:17:51	U	SMPL
P4693-01 - 1	BP-G5-WC	2409	0.5304	-	7471B	PPB	06 Nov 2024 11:33:36	U	SMPL
P4693-05 - 1	BP-G4-WC	444	0.0953	-	7471B	PPB	06 Nov 2024 11:35:53	U	SMPL
P4694-01 - 1	Z-03A	14563	3.2220	-	7471B	PPB	06 Nov 2024 11:38:10	U	SMPL
CCV33 - 1	CCV33	24815	5.4923	-	7471B	PPB	06 Nov 2024 11:40:25	U	SMPL
CCB33 - 1	CCB33	-260	-0.0606	-	7471B	PPB	06 Nov 2024 11:42:43	U	SMPL
P4694-05 - 1	Z-04	13151	2.9093	-	7471B	PPB	06 Nov 2024 11:45:01	U	SMPL
P4695-01 - 1	Z-01	20483	4.5330	-	7471B	PPB	06 Nov 2024 11:47:16	U	SMPL
P4697-01 - 1	TP-1	877	0.1912	-	7471B	PPB	06 Nov 2024 11:49:34	U	SMPL
P4700-01 - 1	MH-8	17328	3.8343	-	7471B	PPB	06 Nov 2024 11:51:52	U	SMPL
P4701-01 - 1	BP-F3	799	0.1739	-	7471B	PPB	06 Nov 2024 11:54:08	U	SMPL
P4701-05 - 1	BP-F4	176	0.0359	-	7471B	PPB	06 Nov 2024 11:56:27	U	SMPL
P4709-01 - 1	HR-02-110424	1460	0.3203	-	7471B	PPB	06 Nov 2024 11:58:43	U	SMPL
P4709-03 - 1	HR-03-110424	2901	0.6394	-	7471B	PPB	06 Nov 2024 12:00:59	U	SMPL
P4711-01 - 1	CF-613-COMP-16	13497	2.9859	-	7471B	PPB	06 Nov 2024 12:03:16	U	SMPL
P4711-06 - 1	CF-613-COMP-17	9436	2.0866	-	7471B	PPB	06 Nov 2024 12:05:32	U	SMPL
CCV34 - 1	CCV34	22547	4.9901	-	7471B	PPB	06 Nov 2024 12:07:50	U	SMPL
CCB34 - 1	CCB34	-217	-0.0511	-	7471B	PPB	06 Nov 2024 12:10:07	U	SMPL
P4711-06DUP - 1	CF-613-COMP-17DUP	9593	2.1214	-	7471B	PPB	06 Nov 2024 12:12:25	U	SMPL
P4711-06MS - 1	CF-613-COMP-17MS	26405	5.8444	-	7471B	PPB	06 Nov 2024 12:14:41	U	SMPL
P4711-06MSD - 1	CF-613-COMP-17MSD	27454	6.0768	-	7471B	PPB	06 Nov 2024 12:16:58	U	SMPL
P4711-06LX5 - 1		1690	0.3712	-	7471B	PPB	06 Nov 2024 12:19:16	U	SMPL
P4711-06A - 1		27431	6.0717	-	7471B	PPB	06 Nov 2024 12:24:16	U	SMPL
CCV35 - 1	CCV35	21176	4.6865	-	7471B	PPB	06 Nov 2024 12:29:26	U	SMPL
CCB35 - 1	CCB35	-235	-0.0551	-	7471B	PPB	06 Nov 2024 12:31:41	U	SMPL