

L3133315


7470A

INSTRUMENT ID: CVD

Linear

μ Abs.:

31883



A= 0.0000e+000

B= 3.1534e-004

C= -6.6424e-002

Rho= 0.9999355

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.0	0.000	-0.070	-0.070	-10	0.000	-10	0				5
0.2	0.200	0.239	0.039	967	0.0 %	967					20
2.5	2.500	2.557	0.057	8319	0.0 %	8319					2
5.0	5.000	4.974	-0.026	15985	0.0 %	15985					1
7.5	7.500	7.512	0.012	24034	0.0 %	24034					0
10.0	10.000	9.988	-0.012	31883	0.0 %	31883					0

LB133315 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Nov 2024 15:19:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	-10	-	0.0000	7470A	PPB	06 Nov 2024 15:27:22	S	Std
0.2 - 1	SD.2	967	-	0.2000	7470A	PPB	06 Nov 2024 15:32:14	S	Std
2.5 - 1	SD.5	8319	-	2.5000	7470A	PPB	06 Nov 2024 15:34:31	S	Std
5.0 - 1	SD	15985	-	5.0000	7470A	PPB	06 Nov 2024 15:36:48	S	Std
7.5 - 1	SD.5	24034	-	7.5000	7470A	PPB	06 Nov 2024 15:44:45	S	Std
10.0 - 1	SD	31883	-	10.0000	7470A	PPB	06 Nov 2024 15:52:22	S	Std
ICV71 - 1	ICV71	13653	4.2389	-	7470A	PPB	06 Nov 2024 15:57:38	U	SMPL
ICB71 - 1	ICB71	-202	-0.1301	-	7470A	PPB	06 Nov 2024 15:59:53	U	SMPL
CCV36 - 1	CCV36	15854	4.9330	-	7470A	PPB	06 Nov 2024 16:02:09	U	SMPL
CCB36 - 1	CCB36	-253	-0.1462	-	7470A	PPB	06 Nov 2024 16:04:24	U	SMPL
CRA - 1	CRA	816	0.1909	-	7470A	PPB	06 Nov 2024 16:06:39	U	SMPL
HighStd - 1	HighStd	31494	9.8648	-	7470A	PPB	06 Nov 2024 16:08:54	U	SMPL
ChkStd - 1	ChkStd	21673	6.7679	-	7470A	PPB	06 Nov 2024 16:11:09	U	SMPL
PB164692BL - 1	PBW	-395	-0.1910	-	7470A	PPB	06 Nov 2024 16:13:27	U	SMPL
PB164692BS - 1	LCSW	11945	3.7003	-	7470A	PPB	06 Nov 2024 16:15:45	U	SMPL
P4693-04 - 1	BP-G5-WC	-272	-0.1522	-	7470A	PPB	06 Nov 2024 16:18:01	U	SMPL
P4693-08 - 1	BP-G4-WC	-91	-0.0951	-	7470A	PPB	06 Nov 2024 16:20:18	U	SMPL
P4694-04 - 1	Z-03A	-106	-0.0999	-	7470A	PPB	06 Nov 2024 16:22:35	U	SMPL
P4694-08 - 1	Z-04	-5	-0.0680	-	7470A	PPB	06 Nov 2024 16:24:52	U	SMPL
P4695-04 - 1	Z-01	-30	-0.0759	-	7470A	PPB	06 Nov 2024 16:27:07	U	SMPL
CCV37 - 1	CCV37	16395	5.1035	-	7470A	PPB	06 Nov 2024 16:29:22	U	SMPL
CCB37 - 1	CCB37	-216	-0.1345	-	7470A	PPB	06 Nov 2024 16:31:36	U	SMPL
P4700-04 - 1	MH-8	-59	-0.0850	-	7470A	PPB	06 Nov 2024 16:33:53	U	SMPL
P4701-04 - 1	BP-F3	14	-0.0620	-	7470A	PPB	06 Nov 2024 16:36:08	U	SMPL
P4701-08 - 1	BP-F4	-51	-0.0825	-	7470A	PPB	06 Nov 2024 16:38:24	U	SMPL
P4711-05 - 1	CF-613-COMP-16	-5	-0.0680	-	7470A	PPB	06 Nov 2024 16:40:39	U	SMPL
P4711-10 - 1	CF-613-COMP-17	42	-0.0532	-	7470A	PPB	06 Nov 2024 16:42:55	U	SMPL
P4711-10DUP - 1	CF-613-COMP-17DUP	-16	-0.0715	-	7470A	PPB	06 Nov 2024 16:45:11	U	SMPL
P4711-10MS - 1	CF-613-COMP-17MS	14623	4.5448	-	7470A	PPB	06 Nov 2024 16:47:27	U	SMPL
P4711-10MSD - 1	CF-613-COMP-17MSD	12397	3.8428	-	7470A	PPB	06 Nov 2024 16:49:45	U	SMPL
CCV38 - 1	CCV38	16674	5.1915	-	7470A	PPB	06 Nov 2024 16:57:17	U	SMPL
CCB38 - 1	CCB38	-241	-0.1424	-	7470A	PPB	06 Nov 2024 16:59:32	U	SMPL
PB164662TB - 1	PB164662TB	8	-0.0639	-	7470A	PPB	06 Nov 2024 17:04:04	U	SMPL
P4711-10LX5 - 1		-12	-0.0702	-	7470A	PPB	06 Nov 2024 17:06:20	U	SMPL
P4711-10A - 1		13226	4.1042	-	7470A	PPB	06 Nov 2024 17:08:35	U	SMPL
CCV39 - 1	CCV39	16773	5.2227	-	7470A	PPB	06 Nov 2024 17:13:07	U	SMPL
CCB39 - 1	CCB39	-306	-0.1629	-	7470A	PPB	06 Nov 2024 17:15:25	U	SMPL