

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133316

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	11/06/2024	08:35
2	P4495-05		1	8	45.00	113.0	112.7	11/06/2024	09:05
3	P4703-01		1	14	100.00	>212.0	>212.0	11/06/2024	09:35
4	P4703-01DUP		1	14	100.00	>212.0	>212.0	11/06/2024	10:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

Lb 133316

WorkList Name : FP-1106 WorkList ID : 185170 Department : Wet-Chemistry Date : 11-06-2024 11:03:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4495-05	PT-FP-SOIL	Solid	Flash Point	Cool 4 deg C	CHEM02	QA Of	10/21/2024	1010B
P4703-01	S0-1	Solid	Flash Point	Cool 4 deg C	PSEG03	L23	11/04/2024	1010B

Date/Time 11/06/2024 08:25
Raw Sample Received by: RM CWG
Raw Sample Relinquished by: RM CWG

Date/Time 11/06/2024 10:45
Raw Sample Received by: RM CWG
Raw Sample Relinquished by: RM CWG