

LB139324


SFAM01.1

INSTRUMENT ID: 101

Linear

 $\mu$  Abs.:

44033



A= 0.0000e+000

B= 2.2918e-004 slope

C= -1.9561e-001 y-intercept

Rho= 0.9997227

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.00   | 0.000  | -0.031 | -0.031 | 719   | 0.000      | 719   |       |       |       |       | 3  |
| 0.05   | 0.050  |        |        |       |            | 442   | 0     |       |       |       | 3  |
| 0.20   | 0.200  | 0.194  | -0.006 | 1698  | 0.0 %      | 1698  |       |       |       |       | 2  |
| 2.50   | 2.500  | 2.547  | 0.047  | 11966 | 0.0 %      | 11966 |       |       |       |       | 2  |
| 5.00   | 5.000  | 4.932  | -0.068 | 22373 | 0.0 %      | 22373 |       |       |       |       | 1  |
| 7.50   | 7.500  | 7.663  | 0.163  | 34291 | 0.0 %      | 34291 |       |       |       |       | 2  |
| 10.0   | 10.000 | 9.896  | -0.104 | 44033 | 0.0 %      | 44033 |       |       |       |       | 1  |

LB133324

## INSTRUMENT ID : CV1

| Sample ID  | Extended ID | $\mu$ Abs. | Conc.     | Std Conc | Method       | Units | Date           | Type |
|------------|-------------|------------|-----------|----------|--------------|-------|----------------|------|
|            | 0 S0        | 719        | -         |          | 0 SFAM01.1   | PPB   | 11/7/2024 8:03 | Std  |
|            | 0.2 S01     | 1698       | -         |          | 0.2 SFAM01.1 | PPB   | 11/7/2024 8:05 | Std  |
|            | 2.5 S02     | 11966      | -         |          | 2.5 SFAM01.1 | PPB   | 11/7/2024 8:07 | Std  |
|            | 5 S03       | 22373      | -         |          | 5 SFAM01.1   | PPB   | 11/7/2024 8:10 | Std  |
|            | 7.5 S04     | 34291      | -         |          | 7.5 SFAM01.1 | PPB   | 11/7/2024 8:12 | Std  |
|            | 10 S05      | 44033      | -         |          | 10 SFAM01.1  | PPB   | 11/7/2024 8:23 | Std  |
| ICV072     | ICV072      | 17285      | 3.7657 -  |          | SFAM01.1     | PPB   | 11/7/2024 8:26 | SMPL |
| ICB072     | ICB072      | 540        | -0.0719 - |          | SFAM01.1     | PPB   | 11/7/2024 8:28 | SMPL |
| CCV040     | CCV040      | 22558      | 4.9742 -  |          | SFAM01.1     | PPB   | 11/7/2024 8:30 | SMPL |
| CCB040     | CCB040      | 498        | -0.0815 - |          | SFAM01.1     | PPB   | 11/7/2024 8:33 | SMPL |
| P4500-01   | MBHCN1      | 7916       | 1.6186 -  |          | SFAM01.1     | PPB   | 11/7/2024 8:37 | SMPL |
| P4500-02   | MBHCN2      | 1463       | 0.1397 -  |          | SFAM01.1     | PPB   | 11/7/2024 8:39 | SMPL |
| P4500-03   | MBHCN3      | 8062       | 1.652 -   |          | SFAM01.1     | PPB   | 11/7/2024 8:42 | SMPL |
| P4500-04   | MBHCN4      | 1923       | 0.2451 -  |          | SFAM01.1     | PPB   | 11/7/2024 8:44 | SMPL |
| P4500-05   | MBHCN5      | 319        | -0.1225 - |          | SFAM01.1     | PPB   | 11/7/2024 8:46 | SMPL |
| P4500-06   | MBHCN6      | 390        | -0.1062 - |          | SFAM01.1     | PPB   | 11/7/2024 8:49 | SMPL |
| P4500-07   | MBHCN7      | 1094       | 0.0551 -  |          | SFAM01.1     | PPB   | 11/7/2024 8:51 | SMPL |
| P4500-08   | MBHCN8      | 1169       | 0.0723 -  |          | SFAM01.1     | PPB   | 11/7/2024 8:53 | SMPL |
| P4500-09   | MBHCN9      | 1577       | 0.1658 -  |          | SFAM01.1     | PPB   | 11/7/2024 8:55 | SMPL |
| P4500-10   | MBHCP0      | 1691       | 0.1919 -  |          | SFAM01.1     | PPB   | 11/7/2024 8:58 | SMPL |
| P4500-11   | MBHCP0D     | 1708       | 0.1958 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:00 | SMPL |
| P4500-12   | MBHCP0S     | 12220      | 2.6049 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:02 | SMPL |
| P4500-13   | MBHCP1      | 1680       | 0.1894 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:04 | SMPL |
| P4500-14   | MBHCP2      | 1743       | 0.2038 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:07 | SMPL |
| P4500-15   | MBHCP3      | 2090       | 0.2834 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:09 | SMPL |
| P4500-16   | MBHCP4      | 5852       | 1.1455 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:11 | SMPL |
| P4500-17   | MBHCP5      | 2512       | 0.3801 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:14 | SMPL |
| P4500-18   | MBHCP6      | 1483       | 0.1443 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:16 | SMPL |
| P4500-19   | MBHCP7      | 779        | -0.0171 - |          | SFAM01.1     | PPB   | 11/7/2024 9:18 | SMPL |
| CCV041     | CCV041      | 23163      | 5.1128 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:20 | SMPL |
| CCB041     | CCB041      | 498        | -0.0815 - |          | SFAM01.1     | PPB   | 11/7/2024 9:23 | SMPL |
| P4500-20   | MBHCX2      | 1736       | 0.2022 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:25 | SMPL |
| P4500-21   | MBHCX3      | 1550       | 0.1596 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:27 | SMPL |
| P4500-22   | MBHCZ6      | 1895       | 0.2387 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:30 | SMPL |
| PB164743BL | PBS743      | 1006       | 0.0349 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:32 | SMPL |
| P4501-01   | MBHCS8      | 2521       | 0.3821 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:34 | SMPL |
| P4501-02   | MBHCS9      | 3320       | 0.5653 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:36 | SMPL |
| P4501-03   | MBHCT0      | 2874       | 0.463 -   |          | SFAM01.1     | PPB   | 11/7/2024 9:39 | SMPL |
| P4501-04   | MBHCT1      | 2390       | 0.3521 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:41 | SMPL |
| P4501-05   | MBHCT2      | 2325       | 0.3372 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:43 | SMPL |
| P4501-06   | MBHCT3      | 1820       | 0.2215 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:45 | SMPL |
| P4501-07   | MBHCT4      | 3952       | 0.7101 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:48 | SMPL |
| P4501-08   | MBHCT5      | 47098      | 10.5981 - |          | SFAM01.1     | PPB   | 11/7/2024 9:50 | SMPL |
| P4501-09   | MBHCT6      | 981        | 0.0292 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:52 | SMPL |
| P4501-10   | MBHCT7      | 1785       | 0.2135 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:54 | SMPL |
| P4501-11   | MBHCT8      | 1917       | 0.2437 -  |          | SFAM01.1     | PPB   | 11/7/2024 9:57 | SMPL |

LB133324

## INSTRUMENT ID : CV1

|            |         |       |           |          |     |                 |      |
|------------|---------|-------|-----------|----------|-----|-----------------|------|
| P4501-12   | MBHCT9  | 2252  | 0.3205 -  | SFAM01.1 | PPB | 11/7/2024 9:59  | SMPL |
| P4501-13   | MBHCW0  | 1667  | 0.1864 -  | SFAM01.1 | PPB | 11/7/2024 10:01 | SMPL |
| P4501-14   | MBHCW0D | 1608  | 0.1729 -  | SFAM01.1 | PPB | 11/7/2024 10:04 | SMPL |
| P4501-15   | MBHCW0S | 13319 | 2.8568 -  | SFAM01.1 | PPB | 11/7/2024 10:06 | SMPL |
| P4501-16   | MBHCW1  | 1199  | 0.0792 -  | SFAM01.1 | PPB | 11/7/2024 10:08 | SMPL |
| CCV042     | CCV042  | 23187 | 5.1183 -  | SFAM01.1 | PPB | 11/7/2024 10:10 | SMPL |
| CCB042     | CCB042  | 620   | -0.0535 - | SFAM01.1 | PPB | 11/7/2024 10:13 | SMPL |
| P4501-17   | MBHCW2  | 2905  | 0.4702 -  | SFAM01.1 | PPB | 11/7/2024 10:15 | SMPL |
| P4501-18   | MBHCW3  | 1755  | 0.2066 -  | SFAM01.1 | PPB | 11/7/2024 10:17 | SMPL |
| P4501-19   | MBHCW4  | 2129  | 0.2923 -  | SFAM01.1 | PPB | 11/7/2024 10:19 | SMPL |
| P4501-20   | MBHCW5  | 2077  | 0.2804 -  | SFAM01.1 | PPB | 11/7/2024 10:22 | SMPL |
| P4501-21   | MBHCW6  | 1858  | 0.2302 -  | SFAM01.1 | PPB | 11/7/2024 10:24 | SMPL |
| P4501-22   | MBHCW7  | 1373  | 0.1191 -  | SFAM01.1 | PPB | 11/7/2024 10:26 | SMPL |
| PB164742BL | PBS742  | 526   | -0.0751 - | SFAM01.1 | PPB | 11/7/2024 10:28 | SMPL |
| CCV043     | CCV043  | 23309 | 5.1463 -  | SFAM01.1 | PPB | 11/7/2024 10:31 | SMPL |
| CCB043     | CCB043  | 715   | -0.0317 - | SFAM01.1 | PPB | 11/7/2024 10:33 | SMPL |