

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB133326

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	11/07/2024	08:20
2	P4662-06		1	11	100.00	>212.0	>212.0	11/07/2024	08:50
3	P4662-06DUP		1	11	100.00	>212.0	>212.0	11/07/2024	09:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133326

WorkList Name : fp-11--24

WorkList ID : 185192

Department : Wet-Chemistry

Date : 11-07-2024 08:06:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4662-06	102524-D	Solid	Flash Point	Cool 4 deg C	PSEG03	K61	10/31/2024	1010B

Date/Time 11/07/2024 08:10
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC

Date/Time 11/07/2024 10:00
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC