

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : sohil

Run Number: LB133327

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/07/2024	10:10
2	CAL2	1	Water	NA	NA	20.3	7.00	11/07/2024	10:11
3	CAL3	1	Water	NA	NA	20.3	10.02	11/07/2024	10:15
4	ICV	1	Water	NA	NA	20.2	6.99	11/07/2024	10:20
5	CCV1	1	Water	NA	NA	20.3	2.01	11/07/2024	10:22
6	P4662-06	1	Solid	20.02	20	23.1	5.17	11/07/2024	10:30
7	P4662-06DUP	1	Solid	20.03	20	23.2	5.18	11/07/2024	10:31
8	CCV2	1	Water	NA	NA	20.3	12.02	11/07/2024	10:35

WORKLIST(Hardcopy Internal Chain)

133327

WorkList Name : ph p4662

WorkList ID : 185206

Department : Wet-Chemistry

Date : 11-07-2024 09:59:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4662-06	102524-D	Solid	pH	Cool 4 deg C	PSEG03	K61	10/31/2024	9045D

Date/Time 11/07/24 10:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11/07/24 12:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]