

PERCENT SOLID

Supervisor: sohil
Analyst: jignesh
Date: 11/8/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 11/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 11/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133343

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P4756-01	BP-B4	1	1.15	8.77	9.92	8.94	88.8	
P4756-02	BP-B4-EPH	2	1.19	8.52	9.71	8.69	88.0	
P4756-03	BP-B4-VOC	3	1.19	8.72	9.91	8.93	88.8	
P4768-01	B-3-1	4	1.16	8.50	9.66	9.17	94.2	
P4768-02	B-3-2	5	1.16	8.62	9.78	9.36	95.1	
P4768-03	B-3-3	6	1.19	8.55	9.74	9.04	91.8	
P4768-04	B-4	7	1.18	8.45	9.63	9.22	95.1	
P4768-05	B-5	8	1.19	8.53	9.72	8.87	90.0	
P4768-06	B-6-2	9	1.15	8.43	9.58	9.1	94.3	
P4768-07	B-6-3	10	1.19	8.44	9.63	9.24	95.4	
P4769-01	PIPE-1	11	1.00	1.00	2.00	2.00	100.0	wipe sample
P4769-02	PIPE-2	12	1.00	1.00	2.00	2.00	100.0	wipe sample
P4769-03	PIPE-3	13	1.00	1.00	2.00	2.00	100.0	wipe sample
P4769-04	PIPE-4	14	1.00	1.00	2.00	2.00	100.0	wipe sample
P4769-05	PIPE-5	15	1.00	1.00	2.00	2.00	100.0	wipe sample
P4776-03	MDL-SOIL-QT4-2024-07	16	1.00	1.00	2.00	2.00	100.0	
P4776-04	MDL-SOIL-QT4-2024-08	17	1.00	1.00	2.00	2.00	100.0	
P4776-05	MDL-MED-SOIL-QT4-2024-07	18	1.00	1.00	2.00	2.00	100.0	
P4776-06	MDL-MED-SOIL-QT4-2024-08	19	1.00	1.00	2.00	2.00	100.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$